

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070123\
 Data File : PR062230.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jun 2023 17:31
 Operator : YP\AJ
 Sample : AR1254ICC200
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 30 18:14:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR070123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 30 18:12:29 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	2.959	24509433	32097677	10.172	10.322
2) SA Decachlor...	9.617	8.241	31193425	77536987	21.200	21.233
Target Compounds						
26) L6 AR-1254-1	5.827	4.923	18477582	37676973	213.557	212.152
27) L6 AR-1254-2	6.064	5.082	27480101	33254875	214.645	212.558
28) L6 AR-1254-3	6.457	5.507	26754456	54113578	210.574	209.057
29) L6 AR-1254-4	6.769	5.754	18110089	34402653	210.160	210.822
30) L6 AR-1254-5	7.225	6.201	19966779	51309196	212.076	210.502



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070123\
Data File : PR062230.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jun 2023 17:31
Operator : YP\AJ
Sample : AR1254ICC200
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 30 18:14:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR070123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 30 18:12:29 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 5.827 min
Delta R.T.: 0.000 min
Response: 18477582
Conc: 213.56 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 4.923 min
Delta R.T.: 0.000 min
Response: 37676973
Conc: 212.15 ng/ml




#27 AR-1254-2

R.T.: 6.064 min
Delta R.T.: 0.000 min
Response: 27480101
Conc: 214.65 ng/ml



#27 AR-1254-2


R.T.: 5.082 min
Delta R.T.: 0.000 min
Response: 33254875
Conc: 212.56 ng/ml



#28 AR-1254-3


R.T.: 6.457 min
Delta R.T.: 0.000 min
Response: 26754456
Conc: 210.57 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 5.507 min
Delta R.T.: 0.000 min
Response: 54113578
Conc: 209.06 ng/ml




#29 AR-1254-4

R.T.: 6.769 min
Delta R.T.: 0.000 min
Response: 18110089
Conc: 210.16 ng/ml



#29 AR-1254-4


R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 34402653
Conc: 210.82 ng/ml



#30 AR-1254-5

R.T.: 7.225 min
Delta R.T.: 0.000 min
Response: 19966779
Conc: 212.08 ng/ml

Instrument :
ECD_R
ClientSampleId :



#30 AR-1254-5

R.T.: 6.201 min
Delta R.T.: 0.000 min
Response: 51309196
Conc: 210.50 ng/ml