

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070318\
 Data File : PR029767.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jun 2018 14:39
 Operator : UA/SM
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Manual Integrations
 APPROVED

Sohil
 7/2/2018 8:36:56 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 29 14:52:59 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062718.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 23 08:41:26 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.565	3.714	428.6E6	839.1E6	18.624m	17.748
2) SA Decachlor...	10.332	8.641	452.2E6	304.0E6	16.153	15.392

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070318\
Data File : PR029767.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jun 2018 14:39
Operator : UA/SM
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Sohil
7/2/2018 8:36:56 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 29 14:52:59 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062718.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 23 08:41:26 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.565 min
Delta R.T.: -0.008 min
Response: 428633006
Conc: 18.62 ng/ml m

Instrument :
ECD_R
ClientSampleId :
I.BLK


Manual Integrations
APPROVED

Sohil
7/2/2018 8:36:56 AM




#1 Tetrachloro-m-xylene

R.T.: 3.714 min
Delta R.T.: 0.001 min
Response: 839055411
Conc: 17.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.332 min
Delta R.T.: -0.018 min
Response: 452220949
Conc: 16.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.641 min
Delta R.T.: -0.016 min
Response: 303981979
Conc: 15.39 ng/ml