

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR070518.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jul 03 19:13:30 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR029859.D 1000 =PR029857.D 750 =PR029858.D
 250 =PR029860.D 50 =PR029861.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	2.650	2.478	2.446	2.568	2.758	2.580 E7	4.94
2) SA	Decachlorobiphenyl	2.476	2.378	2.491	2.558	3.010	2.583 E7	9.58
3) L1	AR-1016-1	5.997	5.546	5.647	5.948	6.388	5.905 E5	5.61
4) L1	AR-1016-2	7.699	7.217	7.171	7.767	8.412	7.653 E5	6.58
5) L1	AR-1016-3	6.542	6.186	6.124	6.519	7.170	6.508 E5	6.38
6) L1	AR-1016-4	6.390	6.023	5.881	6.236	7.060	6.318 E5	7.26
7) L1	AR-1016-5	6.849	6.545	6.440	6.676	7.204	6.743 E5	4.44
8) L2	AR-1221-1	3.212					3.212 E5	0.00
9) L2	AR-1221-2	2.386					2.386 E5	0.00
10) L2	AR-1221-3	7.691					7.691 E5	0.00
11) L3	AR-1232-1	4.668					4.668 E5	0.00
12) L3	AR-1232-2	2.526					2.526 E5	0.00
13) L3	AR-1232-3	3.257					3.257 E5	0.00
14) L3	AR-1232-4	2.678					2.678 E5	0.00
15) L3	AR-1232-5	2.742					2.742 E5	0.00
16) L4	AR-1242-1	4.574					4.574 E5	0.00
17) L4	AR-1242-2	6.856					6.856 E5	0.00
18) L4	AR-1242-3	5.796					5.796 E5	0.00
19) L4	AR-1242-4	5.101					5.101 E5	0.00
20) L4	AR-1242-5	5.031					5.031 E5	0.00
21) L5	AR-1248-1	8.765					8.765 E5	0.00
22) L5	AR-1248-2	9.378					9.378 E5	0.00
23) L5	AR-1248-3	1.985					1.985 E6	0.00
24) L5	AR-1248-4	9.836					9.836 E5	0.00
25) L5	AR-1248-5	1.075					1.075 E6	0.00
26) L6	AR-1254-1	5.483					5.483 E5	0.00
27) L6	AR-1254-2	1.561					1.561 E6	0.00
28) L6	AR-1254-3	1.711					1.711 E6	0.00
29) L6	AR-1254-4	1.339					1.339 E6	0.00
30) L6	AR-1254-5	1.581					1.581 E6	0.00
31) L7	AR-1260-1	1.267	1.183	1.184	1.292	1.393	1.264 E6	6.91
32) L7	AR-1260-2	1.590	1.506	1.499	1.594	1.875	1.613 E6	9.51
33) L7	AR-1260-3	1.223	1.170	1.162	1.214	1.378	1.229 E6	7.10
34) L7	AR-1260-4	2.839	2.768	2.756	2.787	3.079	2.846 E6	4.73
35) L7	AR-1260-5	1.662	1.597	1.611	1.659	1.901	1.686 E6	7.32
36) L8	AR-1262-1	1.027					1.027 E6	0.00
37) L8	AR-1262-2	1.377					1.377 E6	0.00
38) L8	AR-1262-3	1.845					1.845 E6	0.00
39) L8	AR-1262-4	1.626					1.626 E6	0.00
40) L8	AR-1262-5	3.509					3.509 E6	0.00
41) L9	AR-1268-1	3.692					3.692 E6	0.00
42) L9	AR-1268-2	3.423					3.423 E6	0.00
43) L9	AR-1268-3	2.897					2.897 E6	0.00
44) L9	AR-1268-4	1.126					1.126 E6	0.00
45) L9	AR-1268-5	8.440					8.440 E6	0.00

Signal #2 Calibration Files

500 =PR029859.D 1000 =PR029857.D 750 =PR029858.D
 250 =PR029860.D 50 =PR029861.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	4.423	4.104	4.139	4.264	4.642	4.314 E7	5.14
2) SA	Decachlorobiphenyl	1.945	1.770	1.862	2.099	2.695	2.075 E7	17.71
3) L1	AR-1016-1	1.344	1.144	1.237	1.379	1.378	1.296 E6	7.95
4) L1	AR-1016-2	1.439	1.281	1.335	1.440	1.651	1.429 E6	9.91
5) L1	AR-1016-3	1.423	1.235	1.312	1.453	1.753	1.435 E6	13.78

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR070518.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jul 03 19:13:30 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR029859.D 1000 =PR029857.D 750 =PR029858.D
 250 =PR029860.D 50 =PR029861.D

	Compound	500	1000	750	250	50	Avg	%RSD
6)	L1 AR-1016-4	1.559	1.368	1.440	1.577	1.909	1.571 E6	13.23
7)	L1 AR-1016-5	1.597	1.450	1.490	1.620	2.196	1.671 E6	18.09
8)	L2 AR-1221-1	6.192					6.192 E5	0.00
9)	L2 AR-1221-2	4.846					4.846 E5	0.00
10)	L2 AR-1221-3	1.536					1.536 E6	0.00
11)	L3 AR-1232-1	9.673					9.673 E5	0.00
12)	L3 AR-1232-2	6.081					6.081 E5	0.00
13)	L3 AR-1232-3	4.335					4.335 E5	0.00
14)	L3 AR-1232-4	7.419					7.419 E5	0.00
15)	L3 AR-1232-5	7.370					7.370 E5	0.00
16)	L4 AR-1242-1	1.029					1.029 E6	0.00
17)	L4 AR-1242-2	2.322					2.322 E6	0.00
18)	L4 AR-1242-3	2.322					2.322 E6	0.00
19)	L4 AR-1242-4	1.093					1.093 E6	0.00
20)	L4 AR-1242-5	1.216					1.216 E6	0.00
21)	L5 AR-1248-1	1.818					1.818 E6	0.00
22)	L5 AR-1248-2	1.883					1.883 E6	0.00
23)	L5 AR-1248-3	2.052					2.052 E6	0.00
24)	L5 AR-1248-4	3.120					3.120 E6	0.00
25)	L5 AR-1248-5	4.051					4.051 E6	0.00
26)	L6 AR-1254-1	1.274					1.274 E6	0.00
27)	L6 AR-1254-2	2.814					2.814 E6	0.00
28)	L6 AR-1254-3	4.206					4.206 E6	0.00
29)	L6 AR-1254-4	2.622					2.622 E6	0.00
30)	L6 AR-1254-5	2.951					2.951 E6	0.00
31)	L7 AR-1260-1	2.684	2.730	2.507	2.789	3.667	2.876 E6	15.82
32)	L7 AR-1260-2	3.196	3.356	2.981	3.350	4.617	3.500 E6	18.37
33)	L7 AR-1260-3	2.882	3.140	2.700	3.005	4.019	3.149 E6	16.27
34)	L7 AR-1260-4	3.731	4.218	3.506	3.993	5.035	4.097 E6	14.38
35)	L7 AR-1260-5	2.326	2.271	2.118	2.539	3.224	2.496 E6	17.41
36)	L8 AR-1262-1	2.426					2.426 E6	0.00
37)	L8 AR-1262-2	3.288					3.288 E6	0.00
38)	L8 AR-1262-3	3.428					3.428 E6	0.00
39)	L8 AR-1262-4	2.371					2.371 E6	0.00
40)	L8 AR-1262-5	4.377					4.377 E6	0.00
41)	L9 AR-1268-1	3.775					3.775 E6	0.00
42)	L9 AR-1268-2	3.403					3.403 E6	0.00
43)	L9 AR-1268-3	2.731					2.731 E6	0.00
44)	L9 AR-1268-4	1.153					1.153 E6	0.00
45)	L9 AR-1268-5	6.370					6.370 E6	0.00

(#) = Out of Range