

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070523\
 Data File : PR062250.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jul 2023 10:12
 Operator : YP\AJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543285

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 06 00:49:54 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR070123CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 01 02:32:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.687	2.958	47352951	63553979	19.447	21.145
2)	SA Decachlor...	9.619	8.241	58838823	144.0E6	39.758	39.393
Target Compounds							
26)	L6 AR-1254-1	5.826	4.922	34426049	71420178	398.363	402.405
27)	L6 AR-1254-2	6.064	5.081	51657718	63132713	402.378	404.278
28)	L6 AR-1254-3	6.457	5.505	51477592	104.2E6	406.387	403.323
29)	L6 AR-1254-4	6.769	5.753	34292438	63740351	402.771	390.565
30)	L6 AR-1254-5	7.225	6.199	38262256	98221477	410.337	401.809

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070523\
Data File : PR062250.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jul 2023 10:12
Operator : YP\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543285

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 06 00:49:54 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR070123CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 01 02:32:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

