

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR070523\  
 Data File : PR062293.D  
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
 Acq On : 05 Jul 2023 20:44  
 Operator : YP\AJ  
 Sample : 03387-14  
 Misc :  
 ALS Vial : 44 Sample Multiplier: 1

Instrument :  
 ECD\_R  
 ClientSampleId :  
 EX7Z8

Integration File signal 1: autoint1.e  
 Integration File signal 2: autoint2.e  
 Quant Time: Jul 06 01:19:44 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR070123CLP.M  
 Quant Title : GC EXTRACTABLES  
 QLast Update : Sat Jul 01 02:32:00 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 1 µl  
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

| Compound                    | RT#1  | RT#2  | Resp#1   | Resp#2    | ng/ml   | ng/ml      |
|-----------------------------|-------|-------|----------|-----------|---------|------------|
| -----                       |       |       |          |           |         |            |
| System Monitoring Compounds |       |       |          |           |         |            |
| 1) SA Tetrachlo...          | 3.687 | 2.958 | 27883766 | 38709495  | 11.451  | 12.879     |
| 2) SA Decachlor...          | 9.634 | 8.233 | 337.1E6  | 15739.8E6 | 227.800 | 4306.822 # |

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Data\PR070523\  
Data File : PR062293.D  
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch  
Acq On : 05 Jul 2023 20:44  
Operator : YP\AJ  
Sample : 03387-14  
Misc :  
ALS Vial : 44 Sample Multiplier: 1

Instrument :  
ECD\_R  
ClientSampleId :  
EX7Z8

Integration File signal 1: autoint1.e  
Integration File signal 2: autoint2.e  
Quant Time: Jul 06 01:19:44 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD\_R\Method\PR070123CLP.M  
Quant Title : GC EXTRACTABLES  
QLast Update : Sat Jul 01 02:32:00 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1 µl  
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2  
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

