

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070822\
 Data File : PR055233.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Jul 2022 23:55
 Operator : AJ\MA
 Sample : AR1248CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12483379

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 09 04:37:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062222CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 23 10:22:07 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.638	2.909	85232922	38726940	21.000	19.785
2) SA Decachlor...	9.530	8.134	63128713	60569010	43.890	40.286
Target Compounds						
21) L5 AR-1248-1	4.860	4.020	28375009	14897387	405.569	381.756
22) L5 AR-1248-2	5.152	4.266	34800714	19977259	401.830	364.950
23) L5 AR-1248-3	5.363	4.307	38852816	20470914	398.422	367.404
24) L5 AR-1248-4	5.795	4.481	36957971	24584283	391.092	361.192
25) L5 AR-1248-5	5.835	4.891	37740269	24966951	401.377	370.993

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR070822\
Data File : PR055233.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Jul 2022 23:55
Operator : AJ\MA
Sample : AR1248CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12483379

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 09 04:37:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR062222CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 23 10:22:07 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

