

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071118\
 Data File : PR030143.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Jul 2018 13:29
 Operator : MA/SM
 Sample : PB110950BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 PB110950BL

Manual Integrations
 APPROVED

Sohil
 7/11/2018 5:49:30 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 10 14:29:41 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 10 01:27:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.573	3.701	298.9E6	463.5E6	19.040m	17.475
2) SA Decachlor...	10.355	8.631	530.5E6	438.1E6	18.395m	19.426

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071118\
Data File : PR030143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Jul 2018 13:29
Operator : MA/SM
Sample : PB110950BL
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
PB110950BL


Manual Integrations
APPROVED

Sohil
7/11/2018 5:49:30 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 10 14:29:41 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071018.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 10 01:27:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.573 min
Delta R.T.: 0.011 min
Response: 298940448
Conc: 19.04 ng/ml m

Instrument :
ECD_R
ClientSampleId :
PB110950BL


Manual Integrations
APPROVED

Sohil
7/11/2018 5:49:30 PM




#1 Tetrachloro-m-xylene

R.T.: 3.701 min
Delta R.T.: -0.007 min
Response: 463452098
Conc: 17.48 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.355 min
Delta R.T.: 0.030 min
Response: 530541458
Conc: 18.39 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.631 min
Delta R.T.: 0.007 min
Response: 438058429
Conc: 19.43 ng/ml