

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR071120CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 11 00:08:00 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046109.D 1600 =PR046111.D 800 =PR046110.D
 200 =PR046108.D 100 =PR046107.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.787	1.606	1.844	1.789	1.740	1.753	E6	5.12
2) SA	Decachlorobiphenyl	1.300	1.197	1.260	1.373	1.408	1.308	E6	6.49
3) L1	AR-1016-1	5.884	5.152	5.837	6.059	6.068	5.800	E4	6.49
4) L1	AR-1016-2	8.186	7.300	8.202	8.495	8.477	8.132	E4	6.00
5) L1	AR-1016-3	5.005	4.277	4.864	5.216	5.316	4.936	E4	8.28
6) L1	AR-1016-4	4.173	3.619	4.064	4.352	4.612	4.164	E4	8.85
7) L1	AR-1016-5	3.973	3.489	3.893	4.062	4.046	3.893	E4	6.04
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	7.024	6.342	6.744	7.385	7.636	7.026	E4	7.29
32) L7	AR-1260-2	8.506	7.731	8.167	8.837	9.207	8.490	E4	6.75
33) L7	AR-1260-3	6.538	5.943	6.276	6.717	6.928	6.480	E4	5.93
34) L7	AR-1260-4	7.616	7.102	7.350	7.918	8.078	7.613	E4	5.26
35) L7	AR-1260-5	1.565	1.504	1.554	1.594	1.641	1.572	E5	3.22
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR046109.D 1600 =PR046111.D 800 =PR046110.D
 200 =PR046108.D 100 =PR046107.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.232	1.098	1.268	1.220	1.173	1.198	E7	5.47
2) SA	Decachlorobiphenyl	0.965	0.890	0.904	1.011	1.030	0.960	E7	6.47
3) L1	AR-1016-1	4.299	3.718	4.232	4.423	4.559	4.246	E5	7.55
4) L1	AR-1016-2	5.855	5.156	5.841	5.921	5.936	5.742	E5	5.75
5) L1	AR-1016-3	2.901	2.610	2.896	2.909	2.818	2.827	E5	4.48

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR071120CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 11 00:08:00 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046109.D 1600 =PR046111.D 800 =PR046110.D
 200 =PR046108.D 100 =PR046107.D

Compound			400	1600	800	200	100	Avg	%RSD

6)	L1	AR-1016-4	2.285	1.967	2.217	2.363	2.376	2.241 E5	7.43
7)	L1	AR-1016-5	3.072	2.704	2.987	3.166	3.157	3.017 E5	6.29
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	5.368	4.937	5.182	5.559	5.726	5.354 E5	5.79
32)	L7	AR-1260-2	6.511	5.975	6.183	6.724	6.868	6.452 E5	5.75
33)	L7	AR-1260-3	6.000	5.638	5.827	6.099	6.189	5.951 E5	3.70
34)	L7	AR-1260-4	5.077	4.791	4.932	5.213	5.224	5.047 E5	3.69
35)	L7	AR-1260-5	1.279	1.231	1.262	1.290	1.289	1.270 E6	1.94
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range