

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071218\
 Data File : PR030207.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jul 2018 18:53
 Operator : MA/SM
 Sample : J3904-03
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 REFURBISHED-DRUM-1

Manual Integrations
 APPROVED

Sohil
 7/12/2018 7:22:57 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 01:40:01 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 10 01:27:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.544	3.754	82474054	344.5E6	5.253m	12.989m#
2) SA Decachlor...	10.348	8.642	258.0E6	148.4E6	8.944	6.581m#

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071218\
Data File : PR030207.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jul 2018 18:53
Operator : MA/SM
Sample : J3904-03
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
REFURBISHED-DRUM-1


Manual Integrations
APPROVED

Sohil
7/12/2018 7:22:57 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 01:40:01 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071018.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 10 01:27:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.544 min
Delta R.T.: -0.018 min
Response: 82474054
Conc: 5.25 ng/ml m

Instrument :
ECD_R
ClientSampleId :
REFURBISHED-DRUM-1


Manual Integrations
APPROVED

Sohil
7/12/2018 7:22:57 PM




#1 Tetrachloro-m-xylene

R.T.: 3.754 min
Delta R.T.: 0.046 min
Response: 344468869
Conc: 12.99 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: 0.023 min
Response: 257968129
Conc: 8.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.642 min
Delta R.T.: 0.018 min
Response: 148407857
Conc: 6.58 ng/ml m