

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071218\
 Data File : PR030215.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jul 2018 21:18
 Operator : MA/SM
 Sample : J3915-05
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 C0AW7

Manual Integrations
 APPROVED

Sohil
 7/12/2018 7:23:07 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 01:44:52 2018
 Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071018.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 10 01:27:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

1) SA Tetrachlo...	4.562	3.704	357.7E6	551.0E6	22.784m	20.775m
2) SA Decachlor...	10.329	8.620	341.8E6	276.7E6	11.849m	12.271m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071218\
Data File : PR030215.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jul 2018 21:18
Operator : MA/SM
Sample : J3915-05
Misc :
ALS Vial : 26 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
C0AW7


Manual Integrations
APPROVED

Sohil
7/12/2018 7:23:07 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 01:44:52 2018
Quant Method : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071018.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 10 01:27:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.562 min
Delta R.T.: 0.000 min
Response: 357727311
Conc: 22.78 ng/ml m

Instrument :
ECD_R
ClientSampleId :
C0AW7


Manual Integrations
APPROVED

Sohil
7/12/2018 7:23:07 PM




#1 Tetrachloro-m-xylene

R.T.: 3.704 min
Delta R.T.: -0.004 min
Response: 550962689
Conc: 20.77 ng/ml m



#2 Decachlorobiphenyl

R.T.: 10.329 min
Delta R.T.: 0.004 min
Response: 341756696
Conc: 11.85 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.620 min
Delta R.T.: -0.003 min
Response: 276715384
Conc: 12.27 ng/ml m