

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071318\
 Data File : PR030232.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2018 19:47
 Operator : MA/SM
 Sample : AR1660ICC800
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 04:19:32 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071318CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 13 04:18:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.559	3.703	1083.7E6	1360.0E6	40.339	39.856
2)	SA Decachlor...	10.327	8.616	1676.8E6	1110.6E6	79.561	79.425
Target Compounds							
3)	L1 AR-1016-1	4.924	4.069	345.7E6	517.9E6	809.664	799.662
4)	L1 AR-1016-2	5.452	4.838	476.8E6	1142.5E6	804.422	800.353
5)	L1 AR-1016-3	5.718	5.205	681.1E6	951.8E6	793.879	796.270
6)	L1 AR-1016-4	5.900	5.549	497.0E6	962.4E6	803.443	796.580
7)	L1 AR-1016-5	6.194	5.693	474.3E6	680.2E6	799.973	795.528
31)	L7 AR-1260-1	7.315	6.217	915.2E6	1570.7E6	800.027	797.422
32)	L7 AR-1260-2	7.569	6.402	1234.7E6	1822.5E6	799.116	794.331
33)	L7 AR-1260-3	7.851	6.554	1447.2E6	1518.0E6	893.908	794.872
34)	L7 AR-1260-4	8.156	7.018	855.1E6	992.4E6	798.484	794.036
35)	L7 AR-1260-5	8.481	7.257	1935.7E6	2193.7E6	798.866	786.033

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071318\
Data File : PR030232.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2018 19:47
Operator : MA/SM
Sample : AR1660ICC800
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 04:19:32 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071318CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 13 04:18:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

