

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071519\
 Data File : PR039313.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2019 11:12
 Operator : SM\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 01:08:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 16 01:08:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.765	4.611	140.3E6	580.8E6	76.224	76.975
2)	SA Decachlor...	8.705	10.350	221.2E6	783.0E6	87.169	88.322
Target Compounds							
3)	L1 AR-1016-1	4.833	5.766	52333707	223.5E6	775.912	799.202
4)	L1 AR-1016-2	4.850	5.789	87949787	325.4E6	794.236	794.991
5)	L1 AR-1016-3	5.024	5.850	44283530	196.9E6	792.476	799.017
6)	L1 AR-1016-4	5.064	5.949	34694984	167.4E6	783.754	804.516
7)	L1 AR-1016-5	5.274	6.238	47339957	162.8E6	789.092	802.535
31)	L7 AR-1260-1	6.289	7.348	101.1E6	316.2E6	808.399	813.417
32)	L7 AR-1260-2	6.473	7.601	129.4E6	390.5E6	810.204	823.604
33)	L7 AR-1260-3	6.625	7.956	120.9E6	303.7E6	822.500	826.711
34)	L7 AR-1260-4	7.090	8.186	107.1E6	366.7E6	827.270	837.776
35)	L7 AR-1260-5	7.328	8.512	276.8E6	789.8E6	843.334	851.063



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR071519\
Data File : PR039313.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2019 11:12
Operator : SM\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 01:08:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR071519.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 16 01:08:11 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.765 min
Delta R.T.: 0.002 min
Response: 140320583
Conc: 76.22 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 4.611 min
Delta R.T.: 0.002 min
Response: 580849618
Conc: 76.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.705 min
Delta R.T.: 0.000 min
Response: 221174139
Conc: 87.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.350 min
Delta R.T.: 0.000 min
Response: 782980693
Conc: 88.32 ng/ml



#3 AR-1016-1


R.T.: 4.833 min
Delta R.T.: 0.000 min
Response: 52333707
Conc: 775.91 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 5.766 min
Delta R.T.: 0.000 min
Response: 223469679
Conc: 799.20 ng/ml




#4 AR-1016-2

R.T.: 4.850 min
Delta R.T.: 0.000 min
Response: 87949787
Conc: 794.24 ng/ml



#4 AR-1016-2


R.T.: 5.789 min
Delta R.T.: 0.000 min
Response: 325372960
Conc: 794.99 ng/ml



#5 AR-1016-3


R.T.: 5.024 min
Delta R.T.: 0.001 min
Response: 44283530
Conc: 792.48 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.850 min
Delta R.T.: 0.000 min
Response: 196943273
Conc: 799.02 ng/ml




#6 AR-1016-4

R.T.: 5.064 min
Delta R.T.: 0.001 min
Response: 34694984
Conc: 783.75 ng/ml



#6 AR-1016-4


R.T.: 5.949 min
Delta R.T.: 0.000 min
Response: 167354820
Conc: 804.52 ng/ml



#7 AR-1016-5


R.T.: 5.274 min
Delta R.T.: 0.000 min
Response: 47339957
Conc: 789.09 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 6.238 min
Delta R.T.: 0.000 min
Response: 162756690
Conc: 802.54 ng/ml




#31 AR-1260-1

R.T.: 6.289 min
Delta R.T.: 0.001 min
Response: 101109216
Conc: 808.40 ng/ml



#31 AR-1260-1


R.T.: 7.348 min
Delta R.T.: 0.000 min
Response: 316195633
Conc: 813.42 ng/ml



#32 AR-1260-2


R.T.: 6.473 min
Delta R.T.: 0.001 min
Response: 129433463
Conc: 810.20 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 7.601 min
Delta R.T.: 0.000 min
Response: 390533629
Conc: 823.60 ng/ml




#33 AR-1260-3

R.T.: 6.625 min
Delta R.T.: 0.000 min
Response: 120871909
Conc: 822.50 ng/ml



#33 AR-1260-3


R.T.: 7.956 min
Delta R.T.: 0.000 min
Response: 303702131
Conc: 826.71 ng/ml



#34 AR-1260-4


R.T.: 7.090 min
Delta R.T.: 0.000 min
Response: 107108006
Conc: 827.27 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 8.186 min
Delta R.T.: 0.000 min
Response: 366721284
Conc: 837.78 ng/ml



#35 AR-1260-5

R.T.: 7.328 min
Delta R.T.: 0.000 min
Response: 276801471
Conc: 843.33 ng/ml



#35 AR-1260-5

R.T.: 8.512 min
Delta R.T.: 0.000 min
Response: 789840435
Conc: 851.06 ng/ml