

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR072218.M
 Title : GC EXTRACTABLES
 Last Update : Sun Jul 22 23:34:46 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR030499.D 1000 =PR030497.D 750 =PR030498.D
 250 =PR030500.D 50 =PR030501.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.560	2.817	3.338	3.335	3.709	3.352 E7	10.09
2) SA	Decachlorobiphenyl	3.324	3.715	3.760	4.202	4.400	3.880 E7	10.97
3) L1	AR-1016-1	8.386	6.634	7.871	8.225	9.203	8.064 E5	11.61
4) L1	AR-1016-2	1.079	0.862	1.024	1.102	1.256	1.065 E6	13.38
5) L1	AR-1016-3	0.919	0.738	0.883	0.926	1.010	0.895 E6	11.12
6) L1	AR-1016-4	0.889	0.722	0.858	0.931	1.024	0.885 E6	12.46
7) L1	AR-1016-5	0.987	0.800	0.940	0.998	1.170	0.979 E6	13.57
8) L2	AR-1221-1	4.775					4.775 E5	0.00
9) L2	AR-1221-2	3.507					3.507 E5	0.00
10) L2	AR-1221-3	1.075					1.075 E6	0.00
11) L3	AR-1232-1	6.162					6.162 E5	0.00
12) L3	AR-1232-2	3.577					3.577 E5	0.00
13) L3	AR-1232-3	4.501					4.501 E5	0.00
14) L3	AR-1232-4	3.690					3.690 E5	0.00
15) L3	AR-1232-5	3.876					3.876 E5	0.00
16) L4	AR-1242-1	6.856					6.856 E5	0.00
17) L4	AR-1242-2	1.035					1.035 E6	0.00
18) L4	AR-1242-3	8.737					8.737 E5	0.00
19) L4	AR-1242-4	7.554					7.554 E5	0.00
20) L4	AR-1242-5	7.295					7.295 E5	0.00
21) L5	AR-1248-1	1.030					1.030 E6	0.00
22) L5	AR-1248-2	1.163					1.163 E6	0.00
23) L5	AR-1248-3	1.033					1.033 E6	0.00
24) L5	AR-1248-4	1.307					1.307 E6	0.00
25) L5	AR-1248-5	1.355					1.355 E6	0.00
26) L6	AR-1254-1	6.659					6.659 E5	0.00
27) L6	AR-1254-2	1.980					1.980 E6	0.00
28) L6	AR-1254-3	2.144					2.144 E6	0.00
29) L6	AR-1254-4	1.831					1.831 E6	0.00
30) L6	AR-1254-5	1.812					1.812 E6	0.00
31) L7	AR-1260-1	1.789	1.519	1.726	1.822	1.964	1.764 E6	9.22
32) L7	AR-1260-2	2.510	2.159	2.442	2.555	2.835	2.500 E6	9.70
33) L7	AR-1260-3	1.989	1.767	1.967	2.071	2.244	2.008 E6	8.62
34) L7	AR-1260-4	1.837	1.684	1.860	1.949	2.028	1.872 E6	6.90
35) L7	AR-1260-5	4.280	4.058	4.447	4.572	4.882	4.448 E6	6.97
36) L8	AR-1262-1	1.470					1.470 E6	0.00
37) L8	AR-1262-2	1.984					1.984 E6	0.00
38) L8	AR-1262-3	2.853					2.853 E6	0.00
39) L8	AR-1262-4	2.235					2.235 E6	0.00
40) L8	AR-1262-5	5.051					5.051 E6	0.00
41) L9	AR-1268-1	5.373					5.373 E6	0.00
42) L9	AR-1268-2	4.694					4.694 E6	0.00
43) L9	AR-1268-3	4.151					4.151 E6	0.00
44) L9	AR-1268-4	1.805					1.805 E6	0.00
45) L9	AR-1268-5	1.222					1.222 E7	0.00

Signal #2 Calibration Files

500 =PR030499.D 1000 =PR030497.D 750 =PR030498.D
 250 =PR030500.D 50 =PR030501.D

	Compound	500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.318	4.258	5.011	4.995	5.603	5.037 E7	9.96
2) SA	Decachlorobiphenyl	1.991	2.204	2.212	2.630	2.926	2.393 E7	15.78
3) L1	AR-1016-1	1.433	1.151	1.344	1.438	1.522	1.378 E6	10.27
4) L1	AR-1016-2	1.566	1.269	1.467	1.544	1.690	1.507 E6	10.31
5) L1	AR-1016-3	1.546	1.244	1.443	1.555	1.719	1.501 E6	11.64

Method Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR072218.M
 Title : GC EXTRACTABLES
 Last Update : Sun Jul 22 23:34:46 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR030499.D 1000 =PR030497.D 750 =PR030498.D
 250 =PR030500.D 50 =PR030501.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.723	1.402	1.621	1.725	1.883	1.671	E6	10.58
7)	L1 AR-1016-5	1.833	1.508	1.736	1.860	1.994	1.786	E6	10.12
8)	L2 AR-1221-1	6.782					6.782	E5	0.00
9)	L2 AR-1221-2	5.101					5.101	E5	0.00
10)	L2 AR-1221-3	1.666					1.666	E6	0.00
11)	L3 AR-1232-1	9.862					9.862	E5	0.00
12)	L3 AR-1232-2	6.181					6.181	E5	0.00
13)	L3 AR-1232-3	4.483					4.483	E5	0.00
14)	L3 AR-1232-4	8.001					8.001	E5	0.00
15)	L3 AR-1232-5	7.242					7.242	E5	0.00
16)	L4 AR-1242-1	1.173					1.173	E6	0.00
17)	L4 AR-1242-2	2.703					2.703	E6	0.00
18)	L4 AR-1242-3	1.602					1.602	E6	0.00
19)	L4 AR-1242-4	1.265					1.265	E6	0.00
20)	L4 AR-1242-5	1.424					1.424	E6	0.00
21)	L5 AR-1248-1	1.725					1.725	E6	0.00
22)	L5 AR-1248-2	1.790					1.790	E6	0.00
23)	L5 AR-1248-3	2.234					2.234	E6	0.00
24)	L5 AR-1248-4	3.385					3.385	E6	0.00
25)	L5 AR-1248-5	2.422					2.422	E6	0.00
26)	L6 AR-1254-1	1.204					1.204	E6	0.00
27)	L6 AR-1254-2	2.963					2.963	E6	0.00
28)	L6 AR-1254-3	4.906					4.906	E6	0.00
29)	L6 AR-1254-4	3.820					3.820	E6	0.00
30)	L6 AR-1254-5	3.877					3.877	E6	0.00
31)	L7 AR-1260-1	3.500	3.011	3.433	3.675	4.265	3.577	E6	12.74
32)	L7 AR-1260-2	4.350	3.807	4.308	4.729	5.808	4.600	E6	16.31
33)	L7 AR-1260-3	3.338	2.976	3.339	3.650	4.165	3.494	E6	12.73
34)	L7 AR-1260-4	2.321	2.126	2.379	2.620	2.972	2.484	E6	13.08
35)	L7 AR-1260-5	4.845	4.671	5.145	5.674	6.524	5.372	E6	13.93
36)	L8 AR-1262-1	2.930					2.930	E6	0.00
37)	L8 AR-1262-2	3.473					3.473	E6	0.00
38)	L8 AR-1262-3	4.497					4.497	E6	0.00
39)	L8 AR-1262-4	3.132					3.132	E6	0.00
40)	L8 AR-1262-5	5.738					5.738	E6	0.00
41)	L9 AR-1268-1	4.684					4.684	E6	0.00
42)	L9 AR-1268-2	3.987					3.987	E6	0.00
43)	L9 AR-1268-3	3.005					3.005	E6	0.00
44)	L9 AR-1268-4	1.276					1.276	E6	0.00
45)	L9 AR-1268-5	7.127					7.127	E6	0.00

(#) = Out of Range