

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072418\
 Data File : PR030648.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Jul 2018 02:19
 Operator : SM\MA
 Sample : AR1242 LOQ
 Misc : AR1242 LOQ WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1242 LOQ

Manual Integrations
 APPROVED

Sohil
 7/26/2018 8:48:29 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 25 04:42:44 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 25 04:22:29 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.553	3.691	378.9E6	566.4E6	16.639	14.774
2)	SA Decachlor...	10.316	8.593	496.6E6	380.9E6	19.205	19.348
Target Compounds							
16)	L4 AR-1242-1	5.446	4.535	37480230	71355671	84.083	80.621
17)	L4 AR-1242-2	5.711	4.771	61303409	170.1E6	91.781	77.606
18)	L4 AR-1242-3	5.796	4.827	54763902	82760730	97.040	62.601 #
19)	L4 AR-1242-4	5.895	5.023	47233101	82743856	98.239	80.790
20)	L4 AR-1242-5	6.188	5.192	39855151	91148580	84.594	80.927m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072418\
Data File : PR030648.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Jul 2018 02:19
Operator : SM\MA
Sample : AR1242 LOQ
Misc : AR1242 LOQ WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1242 LOQ

Manual Integrations
APPROVED

Sohil
7/26/2018 8:48:29 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 25 04:42:44 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072418.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 25 04:22:29 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

