

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR072620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 25 04:29:05 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046447.D 1600 =PR046449.D 800 =PR046448.D
 200 =PR046446.D 100 =PR046445.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	5.395	4.403	3.784	5.102	3.342	4.405 E5	19.59
2) SA	Decachlorobiphenyl	1.517	1.399	1.351	1.639	1.532	1.488 E6	7.70
3) L1	AR-1016-1	2.917	2.472	2.189	3.061	2.242	2.576 E4	15.34
4) L1	AR-1016-2	4.085	3.509	3.120	4.434	3.223	3.674 E4	15.42
5) L1	AR-1016-3	2.559	2.190	1.971	2.733	2.101	2.311 E4	13.92
6) L1	AR-1016-4	2.126	1.856	1.662	2.323	1.859	1.965 E4	13.20
7) L1	AR-1016-5	2.313	2.014	1.798	2.581	1.592	2.060 E4	19.19
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1	5.696	4.994	4.576	6.409	5.476	5.430 E4	12.86
32) L7	AR-1260-2	7.182	6.407	5.883	7.981	6.782	6.847 E4	11.60
33) L7	AR-1260-3	5.896	5.267	4.848	6.401	5.422	5.567 E4	10.75
34) L7	AR-1260-4	7.075	6.413	5.932	7.678	6.512	6.722 E4	9.99
35) L7	AR-1260-5	1.533	1.462	1.328	1.635	1.449	1.481 E5	7.64
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

Signal #2 Calibration Files

400 =PR046447.D 1600 =PR046449.D 800 =PR046448.D
 200 =PR046446.D 100 =PR046445.D

	Compound	400	1600	800	200	100	Avg	%RSD
1) SA	Tetrachloro-m-xylene	3.514	2.892	2.491	3.416	2.404	2.943 E6	17.38
2) SA	Decachlorobiphenyl	1.126	1.031	1.004	1.219	1.160	1.108 E7	8.10
3) L1	AR-1016-1	1.983	1.667	1.505	2.165	1.772	1.818 E5	14.29
4) L1	AR-1016-2	2.736	2.351	2.086	2.918	2.274	2.473 E5	13.88
5) L1	AR-1016-3	1.303	1.201	1.056	1.414	1.015	1.198 E5	13.91
6) L1	AR-1016-4	1.156	0.992	0.910	1.324	1.169	1.110 E5	14.63

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR072620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 25 04:29:05 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046447.D 1600 =PR046449.D 800 =PR046448.D
 200 =PR046446.D 100 =PR046445.D

Compound			400	1600	800	200	100	Avg	%RSD
7)	L1	AR-1016-5	1.696	1.465	1.330	1.852	1.516	1.572 E5	12.97
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	4.221	3.760	3.439	4.706	4.081	4.041 E5	11.85
32)	L7	AR-1260-2	5.548	4.972	4.576	6.123	5.378	5.319 E5	11.02
33)	L7	AR-1260-3	5.108	4.681	4.259	5.597	4.856	4.900 E5	10.15
34)	L7	AR-1260-4	4.785	4.397	4.049	5.174	4.483	4.578 E5	9.27
35)	L7	AR-1260-5	1.315	1.236	1.140	1.392	1.218	1.260 E6	7.66
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range