

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR072620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 25 07:01:10 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046477.D 1600 =PR046479.D 800 =PR046478.D
 200 =PR046476.D 100 =PR046475.D

	Compound	400	1600	800	200	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	5.225	5.456	5.089	5.659	5.132	5.312	E5	4.53
2) SA	Decachlorobiphenyl	1.654	1.531	1.600	1.788	1.750	1.665	E6	6.33
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	3.915	3.633	3.782	4.328	4.040	3.940	E4	6.72
37) L8	AR-1262-2	1.940	1.906	1.934	2.098	2.026	1.981	E5	4.01
38) L8	AR-1262-3	8.558	8.034	8.398	9.586	8.858	8.687	E4	6.72
39) L8	AR-1262-4	1.028	0.956	0.997	1.110	1.048	1.028	E5	5.60
40) L8	AR-1262-5	8.255	7.728	8.052	8.951	8.597	8.317	E4	5.71
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR046477.D 1600 =PR046479.D 800 =PR046478.D
 200 =PR046476.D 100 =PR046475.D

	Compound	400	1600	800	200	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.676	3.782	3.547	4.114	3.736	3.771	E6	5.59
2) SA	Decachlorobiphenyl	1.300	1.196	1.260	1.421	1.412	1.318	E7	7.39
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR072620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Sat Jul 25 07:01:10 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046477.D 1600 =PR046479.D 800 =PR046478.D
 200 =PR046476.D 100 =PR046475.D

Compound			400	1600	800	200	100	Avg	%RSD
7)	L1	AR-1016-5						0.000	-1.00
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1						0.000	-1.00
32)	L7	AR-1260-2						0.000	-1.00
33)	L7	AR-1260-3						0.000	-1.00
34)	L7	AR-1260-4						0.000	-1.00
35)	L7	AR-1260-5						0.000	-1.00
36)	L8	AR-1262-1	4.146	3.881	4.024	4.648	4.542	4.248 E5	7.82
37)	L8	AR-1262-2	1.703	1.633	1.688	1.842	1.744	1.722 E6	4.52
38)	L8	AR-1262-3	6.859	6.427	6.719	7.581	7.332	6.984 E5	6.69
39)	L8	AR-1262-4	1.304	1.246	1.290	1.410	1.347	1.319 E6	4.71
40)	L8	AR-1262-5	6.996	6.565	6.848	7.576	7.424	7.082 E5	5.87
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

(#) = Out of Range