

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F111921AL\
 Data File : FC056922.D
 Signal(s) : FID1A.CH
 Acq On : 19 Nov 2021 1:34
 Operator : AJ\MA
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_R
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Nov 19 02:07:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 111921.M
 Quant Title : GC Extractables
 QLast Update : Fri Nov 19 00:52:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	101.332	99.687 E3	1.6	98	0.00
2 T n-Decane (C10)	100.139	99.222 E3	0.9	99	0.00
3 T A~Naphthalene (C11.7)	111.662	110.754 E3	0.8	99	0.00
4 T n-Dodecane (C12)	101.392	100.114 E3	1.3	99	0.00
5 T A~2-methylnaphthalene (C12.	110.261	111.191 E3	-0.8	98	0.00
6 T n-Tetradecane (C14)	101.320	99.946 E3	1.4	99	0.00
7 T n-Hexadecane (C16)	102.617	102.024 E3	0.6	99	0.00
8 T n-Octadecane (C18)	107.138	105.299 E3	1.7	99	0.00
9 S ortho-Terphenyl (SURR)	116.047	114.072 E3	1.7	99	0.00
10 T n-Eicosane (C20)	106.955	105.280 E3	1.6	99	0.00
11 T n-Heneicosane (C21)	108.293	106.307 E3	1.8	99	0.00
12 S 1-chlorooctadecane (SURR)	91.038	89.703 E3	1.5	99	0.00
13 T n-Docosane (C22)	107.134	105.683 E3	1.4	99	0.00
14 T n-Tetracosane (C24)	108.256	106.710 E3	1.4	99	0.00
15 T n-Hexacosane (C26)	106.518	105.505 E3	1.0	100	0.00
16 T n-Octacosane (C28)	106.666	105.264 E3	1.3	99	0.00
17 T n-Tricontane (C30)	106.586	104.458 E3	2.0	99	0.00
18 T n-Dotriacontane (C32)	105.480	102.753 E3	2.6	99	0.00
19 T n-Tetratriacontane (C34)	103.887	101.342 E3	2.4	99	0.00
20 T n-Hexatriacontane (C36)	101.007	95.470 E3	5.5	94	0.00
21 T n-Octatriacontane (C38)	91.796	89.371 E3	2.6	98	0.00
22 T n-Tetracontane (C40)	90.084	87.944 E3	2.4	98	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0