

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072723\
 Data File : PR062434.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2023 20:54
 Operator : YP\AJ
 Sample : AR1248ICC400
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12483218

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 03:12:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072723CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 27 03:10:47 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.685	2.956	39956333	61332164	20.000	20.000
2)	SA Decachlor...	9.616	8.236	62553370	156.2E6	40.000	40.000
Target Compounds							
21)	L5 AR-1248-1	4.914	4.080	16754497	27085556	400.000	400.000
22)	L5 AR-1248-2	5.208	4.329	25091457	38686522	400.000	400.000
23)	L5 AR-1248-3	5.421	4.371	28886570	39830537	400.000	400.000
24)	L5 AR-1248-4	5.855	4.548	29767249	48019048	400.000	400.000
25)	L5 AR-1248-5	5.895	4.962	28910695	48140295	400.000	400.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072723\
Data File : PR062434.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2023 20:54
Operator : YP\AJ
Sample : AR1248ICC400
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12483218

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 03:12:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072723CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 27 03:10:47 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

