

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072821\
 Data File : PR051427.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2021 17:29
 Operator : AJ\MA
 Sample : M2940-09
 Misc : AR1232 MDL 25 PPB
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-WATER-03-QT3-2021

Manual Integrations
 APPROVED

mohammad
 7/29/2021 1:10:33 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 04:41:55 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 28 02:51:26 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.628	3.809	147.2E6	136.5E6	18.011	18.360
2) SA Decachlor...	10.502	8.831	187.1E6	139.6E6	19.481	17.457
Target Compounds						
11) L3 AR-1232-1	5.008	4.183	5621890	4204686	31.138	26.039
12) L3 AR-1232-2	5.542	4.913	3031239	4751063	31.182	28.342m
13) L3 AR-1232-3	5.834	5.088	5663345	2157384	27.091	24.756
14) L3 AR-1232-4	5.997	5.170	3213724	1911449	31.125	25.986
15) L3 AR-1232-5	6.082	5.342	1987593	2176114	26.463	26.745

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR072821\
Data File : PR051427.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2021 17:29
Operator : AJ\MA
Sample : M2940-09
Misc : AR1232 MDL 25 PPB
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-03-QT3-2021

Manual Integrations
APPROVED

mohammad
7/29/2021 1:10:33 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 04:41:55 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072721.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 28 02:51:26 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

