

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR073119\
 Data File : PR039878.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jul 2019 10:06
 Operator : SM\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 01:23:57 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 22 14:03:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.750	4.595	40365980	178.6E6	16.819	17.812
2)	SA Decachlor...	8.686	10.326	35125603	144.4E6	14.137	15.378
Target Compounds							
5)	L1 AR-1016-3	5.061	0.000	148827	0	2.143	N.D. #
6)	L1 AR-1016-4	5.061	0.000	148827	0	2.644	N.D. #
9)	L2 AR-1221-2	4.043	0.000	584259	0	24.013	N.D. #
13)	L3 AR-1232-3	5.061	0.000	148827	0	3.521	N.D. #
14)	L3 AR-1232-4	5.069	0.000	174722	0	4.390	N.D. #
18)	L4 AR-1242-3	5.061	0.000	148827	0	2.106	N.D. #
19)	L4 AR-1242-4	5.069	0.000	174722	0	2.498	N.D. #
22)	L5 AR-1248-2	5.061	0.000	148827	0	1.902	N.D. #
23)	L5 AR-1248-3	5.069	0.000	174722	0	1.636	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR073119\
Data File : PR039878.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jul 2019 10:06
Operator : SM\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 01:23:57 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072219.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 22 14:03:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.750 min
Delta R.T.: -0.005 min
Response: 40365980
Conc: 16.82 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 4.595 min
Delta R.T.: -0.007 min
Response: 178554505
Conc: 17.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.686 min
Delta R.T.: -0.009 min
Response: 35125603
Conc: 14.14 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.326 min
Delta R.T.: -0.010 min
Response: 144368153
Conc: 15.38 ng/ml



#5 AR-1016-3


R.T.: 5.061 min
Delta R.T.: 0.046 min
Response: 148827
Conc: 2.14 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 5.842 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.061 min
Delta R.T.: 0.006 min
Response: 148827
Conc: 2.64 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 5.941 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.043 min
Delta R.T.: -0.007 min
Response: 584259
Conc: 24.01 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.893 min
Response: 0
Conc: N.D.




#13 AR-1232-3

R.T.: 5.061 min
Delta R.T.: 0.047 min
Response: 148827
Conc: 3.52 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T. : 5.780 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 5.069 min
Delta R.T.: -0.026 min
Response: 174722
Conc: 4.39 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T. : 5.940 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.061 min
Delta R.T.: 0.045 min
Response: 148827
Conc: 2.11 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 5.842 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.069 min
Delta R.T.: -0.028 min
Response: 174722
Conc: 2.50 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 5.941 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.061 min
Delta R.T.: 0.005 min
Response: 148827
Conc: 1.90 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 6.025 min
Response: 0
Conc: N.D.



#23 AR-1248-3

R.T.: 5.069 min
Delta R.T.: -0.027 min
Response: 174722
Conc: 1.64 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK



#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 6.229 min
Response: 0
Conc: N.D.