

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080124\
 Data File : PR067930.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2024 06:20
 Operator : AJ\MA
 Sample : AR1254ICC800
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12544239

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 08:14:49 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 01 08:09:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.678	3.014	103.3E6	191.1E6	38.121	37.082
2)	SA Decachlor...	9.592	8.337	159.2E6	187.1E6	73.172	72.385
Target Compounds							
26)	L6 AR-1254-1	5.810	4.993	53510174	192.2E6	745.118	718.418
27)	L6 AR-1254-2	6.047	5.155	84329011	169.7E6	727.374	716.298
28)	L6 AR-1254-3	6.440	5.581	103.3E6	272.2E6	734.391	726.545
29)	L6 AR-1254-4	6.752	5.829	74830784	163.9E6	734.144	731.351
30)	L6 AR-1254-5	7.207	6.279	88441326	250.5E6	729.817	729.183

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080124\
Data File : PR067930.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2024 06:20
Operator : AJ\MA
Sample : AR1254ICC800
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12544239

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 08:14:49 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 01 08:09:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

