

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080124\
 Data File : PR067931.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2024 06:34
 Operator : AJ\MA
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12545240

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 01 08:15:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 01 08:09:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.678	3.014	196.6E6	337.6E6	72.566	65.485
2)	SA Decachlor...	9.593	8.338	296.1E6	339.7E6	136.086	131.472
Target Compounds							
26)	L6 AR-1254-1	5.810	4.993	96807282	342.3E6	1348.020	1279.376
27)	L6 AR-1254-2	6.047	5.154	152.9E6	303.6E6	1319.192	1281.678
28)	L6 AR-1254-3	6.439	5.581	192.9E6	487.1E6	1370.761	1299.971
29)	L6 AR-1254-4	6.751	5.829	137.7E6	297.6E6	1351.199	1327.972
30)	L6 AR-1254-5	7.206	6.279	164.2E6	451.5E6	1354.959	1314.082

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080124\
Data File : PR067931.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2024 06:34
Operator : AJ\MA
Sample : AR1254ICC1600
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12545240

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 01 08:15:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 01 08:09:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

