

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080219\
 Data File : PR039975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2019 20:35
 Operator : SM\AJ
 Sample : K4137-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 S15-07-S

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 04:19:12 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 22 14:03:19 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.752	4.596	45802976	174.7E6	19.085	17.429
2) SA Decachlor...	8.684	10.323	52562955	209.7E6	21.155	22.340
Target Compounds						
26) L6 AR-1254-1	5.605	6.593	10750797	16501771	69.706	45.304 #
27) L6 AR-1254-2	5.749	6.807	12919389	75484334	98.672	132.154 #
28) L6 AR-1254-3	6.146	7.175	33477215	62161170	149.834	99.698 #
29) L6 AR-1254-4	6.370	7.455	23429874	70048504	161.122	154.669
30) L6 AR-1254-5	6.782	7.872	60156270	266.2E6	315.920	560.705 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080219\
Data File : PR039975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2019 20:35
Operator : SM\AJ
Sample : K4137-03
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
S15-07-S

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 04:19:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR072219.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 22 14:03:19 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.752 min
Delta R.T.: -0.003 min
Response: 45802976
Conc: 19.08 ng/ml

Instrument :
ECD_R
ClientSampleId :
S15-07-S




#1 Tetrachloro-m-xylene

R.T.: 4.596 min
Delta R.T.: -0.006 min
Response: 174715948
Conc: 17.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.684 min
Delta R.T.: -0.011 min
Response: 52562955
Conc: 21.16 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.323 min
Delta R.T.: -0.013 min
Response: 209721694
Conc: 22.34 ng/ml



#26 AR-1254-1


R.T.: 5.605 min
Delta R.T.: -0.005 min
Response: 10750797
Conc: 69.71 ng/ml

Instrument :
ECD_R
ClientSampleId :
S15-07-S




#26 AR-1254-1

R.T.: 6.593 min
Delta R.T.: -0.008 min
Response: 16501771
Conc: 45.30 ng/ml




#27 AR-1254-2

R.T.: 5.749 min
Delta R.T.: -0.005 min
Response: 12919389
Conc: 98.67 ng/ml



#27 AR-1254-2


R.T.: 6.807 min
Delta R.T.: -0.009 min
Response: 75484334
Conc: 132.15 ng/ml



#28 AR-1254-3


R.T.: 6.146 min
Delta R.T.: -0.006 min
Response: 33477215
Conc: 149.83 ng/ml

Instrument :
ECD_R
ClientSampleId :
S15-07-S




#28 AR-1254-3

R.T.: 7.175 min
Delta R.T.: -0.006 min
Response: 62161170
Conc: 99.70 ng/ml




#29 AR-1254-4

R.T.: 6.370 min
Delta R.T.: -0.006 min
Response: 23429874
Conc: 161.12 ng/ml



#29 AR-1254-4


R.T.: 7.455 min
Delta R.T.: -0.008 min
Response: 70048504
Conc: 154.67 ng/ml



#30 AR-1254-5

R.T.: 6.782 min
Delta R.T.: -0.007 min
Response: 60156270
Conc: 315.92 ng/ml

Instrument :
ECD_R
ClientSampleId :
S15-07-S



#30 AR-1254-5

R.T.: 7.872 min
Delta R.T.: -0.005 min
Response: 266224292
Conc: 560.71 ng/ml