

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080520-8011-504.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 05 09:47:51 2020
 Response Via : Initial Calibration

Calibration Files

0.5 =PR046682.D 0.25 =PR046683.D 0.1 =PR046684.D
 0.05 =PR046685.D .025 =PR046686.D

Compound		0.5	0.25	0.1	0.05	.025	Avg	%RSD	
1) T	EDB	3.428	3.277	2.996	3.315	3.471	3.297	E7	5.64
2) T	DBCP	7.278	7.206	7.020	8.308	7.971	7.557	E7	7.32

Signal #2 Calibration Files

0.5 =PR046682.D 0.25 =PR046683.D 0.1 =PR046684.D
 0.05 =PR046685.D .025 =PR046686.D

Compound		0.5	0.25	0.1	0.05	.025	Avg	%RSD	
1) SA	EDB	1.717	1.684	1.577	1.768	1.706	1.690	E8	4.16
2) SA	DBCP	3.362	3.392	3.164	3.358	3.293	3.314	E8	2.75

(#) = Out of Range