

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080522\
 Data File : PR055639.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 11:30
 Operator : AJ\MA
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 12:02:56 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:01:38 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.635	2.906	328.5E6	158.7E6	65.094	65.119
2)	SA Decachlor...	9.530	8.124	200.6E6	216.9E6	65.872	67.212
Target Compounds							
41)	L9 AR-1268-1	8.130	6.988	165.5E6	190.0E6	658.596	659.226
42)	L9 AR-1268-2	8.220	7.058	151.4E6	173.2E6	655.601	661.560
43)	L9 AR-1268-3	8.437	7.277	126.9E6	145.8E6	655.619	660.963
44)	L9 AR-1268-4	8.840	7.592	58548130	65033035	654.977	664.465
45)	L9 AR-1268-5	9.223	7.886	418.8E6	480.0E6	669.267	676.074



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080522\
Data File : PR055639.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 11:30
Operator : AJ\MA
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 12:02:56 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:01:38 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.635 min
Delta R.T.: 0.000 min
Response: 328510370
Conc: 65.09 ng/ml

Instrument :
ECD_R
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.906 min
Delta R.T.: 0.000 min
Response: 158693201
Conc: 65.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.530 min
Delta R.T.: 0.000 min
Response: 200550204
Conc: 65.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.124 min
Delta R.T.: 0.000 min
Response: 216862734
Conc: 67.21 ng/ml



#41 AR-1268-1


R.T.: 8.130 min
Delta R.T.: 0.000 min
Response: 165455686
Conc: 658.60 ng/ml

Instrument :
ECD_R
ClientSampleId :




#41 AR-1268-1

R.T.: 6.988 min
Delta R.T.: 0.000 min
Response: 189993407
Conc: 659.23 ng/ml




#42 AR-1268-2

R.T.: 8.220 min
Delta R.T.: 0.000 min
Response: 151354250
Conc: 655.60 ng/ml



#42 AR-1268-2


R.T.: 7.058 min
Delta R.T.: 0.000 min
Response: 173176458
Conc: 661.56 ng/ml



#43 AR-1268-3


R.T.: 8.437 min
Delta R.T.: 0.000 min
Response: 126861731
Conc: 655.62 ng/ml

Instrument :
ECD_R
ClientSampleId :




#43 AR-1268-3

R.T.: 7.277 min
Delta R.T.: 0.000 min
Response: 145801055
Conc: 660.96 ng/ml




#44 AR-1268-4

R.T.: 8.840 min
Delta R.T.: 0.000 min
Response: 58548130
Conc: 654.98 ng/ml



#44 AR-1268-4


R.T.: 7.592 min
Delta R.T.: 0.000 min
Response: 65033035
Conc: 664.47 ng/ml



#45 AR-1268-5

R.T.: 9.223 min
Delta R.T.: 0.001 min
Response: 418776752
Conc: 669.27 ng/ml

Instrument :
ECD_R
ClientSampleId :



#45 AR-1268-5

R.T.: 7.886 min
Delta R.T.: 0.001 min
Response: 480018348
Conc: 676.07 ng/ml