

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 07 03:23:01 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046701.D 1600 =PR046703.D 800 =PR046702.D
 200 =PR046700.D 100 =PR046699.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.573	1.678	1.594	1.494	1.738	1.616	E6	5.88
2) SA	Decachlorobiphenyl	2.132	1.957	2.049	2.188	2.237	2.113	E6	5.29
3) L1	AR-1016-1	6.408	6.328	6.328	6.287	7.445	6.559	E4	7.58
4) L1	AR-1016-2	0.906	0.910	0.892	0.881	1.047	0.927	E5	7.34
5) L1	AR-1016-3	5.515	5.308	5.362	5.505	6.562	5.650	E4	9.16
6) L1	AR-1016-4	4.586	4.493	4.492	4.559	5.539	4.734	E4	9.55
7) L1	AR-1016-5	4.727	4.446	4.519	4.443	5.207	4.669	E4	6.91
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1	0.938	0.897	0.908	0.956	1.105	0.961	E5	8.74
32) L7	AR-1260-2	1.162	1.130	1.133	1.175	1.325	1.185	E5	6.82
33) L7	AR-1260-3	0.929	0.887	0.895	0.929	1.017	0.932	E5	5.50
34) L7	AR-1260-4	1.091	1.067	1.064	1.086	1.176	1.097	E5	4.18
35) L7	AR-1260-5	2.319	2.336	2.323	2.299	2.489	2.353	E5	3.27
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR046701.D 1600 =PR046703.D 800 =PR046702.D
 200 =PR046700.D 100 =PR046699.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	8.187	8.823	8.351	7.641	8.583	8.317	E6	5.38
2) SA	Decachlorobiphenyl	1.420	1.338	1.404	1.451	1.485	1.420	E7	3.88
3) L1	AR-1016-1	3.459	3.362	3.380	3.299	4.038	3.508	E5	8.61
4) L1	AR-1016-2	4.792	4.898	4.820	4.689	5.306	4.901	E5	4.87
5) L1	AR-1016-3	1.947	2.163	2.106	1.874	2.137	2.045	E5	6.24

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 07 03:23:01 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046701.D 1600 =PR046703.D 800 =PR046702.D
 200 =PR046700.D 100 =PR046699.D

Compound			400	1600	800	200	100	Avg	%RSD

6)	L1	AR-1016-4	1.677	1.685	1.707	1.715	2.194	1.795 E5	12.44
7)	L1	AR-1016-5	2.645	2.558	2.562	2.566	2.964	2.659 E5	6.55
8)	L2	AR-1221-1						0.000	-1.00
9)	L2	AR-1221-2						0.000	-1.00
10)	L2	AR-1221-3						0.000	-1.00
11)	L3	AR-1232-1						0.000	-1.00
12)	L3	AR-1232-2						0.000	-1.00
13)	L3	AR-1232-3						0.000	-1.00
14)	L3	AR-1232-4						0.000	-1.00
15)	L3	AR-1232-5						0.000	-1.00
16)	L4	AR-1242-1						0.000	-1.00
17)	L4	AR-1242-2						0.000	-1.00
18)	L4	AR-1242-3						0.000	-1.00
19)	L4	AR-1242-4						0.000	-1.00
20)	L4	AR-1242-5						0.000	-1.00
21)	L5	AR-1248-1						0.000	-1.00
22)	L5	AR-1248-2						0.000	-1.00
23)	L5	AR-1248-3						0.000	-1.00
24)	L5	AR-1248-4						0.000	-1.00
25)	L5	AR-1248-5						0.000	-1.00
26)	L6	AR-1254-1						0.000	-1.00
27)	L6	AR-1254-2						0.000	-1.00
28)	L6	AR-1254-3						0.000	-1.00
29)	L6	AR-1254-4						0.000	-1.00
30)	L6	AR-1254-5						0.000	-1.00
31)	L7	AR-1260-1	5.171	5.121	5.124	5.142	6.147	5.341 E5	8.44
32)	L7	AR-1260-2	6.621	6.576	6.595	6.643	7.416	6.770 E5	5.35
33)	L7	AR-1260-3	6.043	6.131	6.092	5.932	6.523	6.144 E5	3.65
34)	L7	AR-1260-4	5.434	5.531	5.501	5.397	5.817	5.536 E5	2.99
35)	L7	AR-1260-5	1.552	1.586	1.590	1.522	1.610	1.572 E6	2.20
36)	L8	AR-1262-1						0.000	-1.00
37)	L8	AR-1262-2						0.000	-1.00
38)	L8	AR-1262-3						0.000	-1.00
39)	L8	AR-1262-4						0.000	-1.00
40)	L8	AR-1262-5						0.000	-1.00
41)	L9	AR-1268-1						0.000	-1.00
42)	L9	AR-1268-2						0.000	-1.00
43)	L9	AR-1268-3						0.000	-1.00
44)	L9	AR-1268-4						0.000	-1.00
45)	L9	AR-1268-5						0.000	-1.00

 (#) = Out of Range