

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 07 05:58:29 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046731.D 1600 =PR046733.D 800 =PR046732.D
 200 =PR046730.D 100 =PR046729.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.208	2.109	2.161	2.203	2.221	2.180	E6	2.10
2)	SA	Decachlorobiphenyl	2.181	1.992	2.077	2.314	2.289	2.171	E6	6.32
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1	6.939	6.269	6.644	7.346	7.313	6.902	E4	6.62
37)	L8	AR-1262-2	3.254	3.099	3.232	3.334	3.331	3.250	E5	2.95
38)	L8	AR-1262-3	1.410	1.282	1.353	1.479	1.566	1.418	E5	7.76
39)	L8	AR-1262-4	1.626	1.493	1.562	1.737	1.732	1.630	E5	6.52
40)	L8	AR-1262-5	1.146	1.060	1.109	1.231	1.206	1.150	E5	6.08
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PR046731.D 1600 =PR046733.D 800 =PR046732.D
 200 =PR046730.D 100 =PR046729.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.176	1.112	1.150	1.160	1.180	1.155	E7	2.34
2) SA	Decachlorobiphenyl	1.585	1.459	1.523	1.678	1.656	1.580	E7	5.79
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR080620CLP.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 07 05:58:29 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046731.D 1600 =PR046733.D 800 =PR046732.D
 200 =PR046730.D 100 =PR046729.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1	5.522	5.015	5.271	5.574	5.765	5.429 E5	5.36
37) L8	AR-1262-2	2.250	2.142	2.227	2.278	2.249	2.229 E6	2.33
38) L8	AR-1262-3	9.035	8.457	8.847	9.378	9.480	9.039 E5	4.58
39) L8	AR-1262-4	1.713	1.625	1.688	1.749	1.739	1.703 E6	2.90
40) L8	AR-1262-5	8.417	7.922	8.206	8.822	8.718	8.417 E5	4.38
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range