

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080622\
 Data File : PR055681.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 22:53
 Operator : AJ\MA
 Sample : N3980-07
 Misc : AR1268 LOD 40 PPB
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 03:58:55 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:39:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.635	2.906	84397607	41402326	19.779	20.621
2) SA Decachlor...	9.526	8.121	42403801	42556525	23.849	23.395
Target Compounds						
41) L9 AR-1268-1	8.128	6.987	11006162	12290813	44.139	43.695
42) L9 AR-1268-2	8.218	7.056	10017676	10117257	43.877	39.546
43) L9 AR-1268-3	8.435	7.275	8567522	8670103	44.490	40.286
44) L9 AR-1268-4	8.836	7.589	3829657	4071707	43.421	42.849
45) L9 AR-1268-5	9.218	7.883	25211285	25815049	41.232	37.885



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080622\
Data File : PR055681.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 22:53
Operator : AJ\MA
Sample : N3980-07
Misc : AR1268 LOD 40 PPB
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT3-2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 03:58:55 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:39:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.635 min
Delta R.T.: -0.001 min
Response: 84397607
Conc: 19.78 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT3-2022




#1 Tetrachloro-m-xylene

R.T.: 2.906 min
Delta R.T.: 0.000 min
Response: 41402326
Conc: 20.62 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.526 min
Delta R.T.: -0.002 min
Response: 42403801
Conc: 23.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.121 min
Delta R.T.: -0.002 min
Response: 42556525
Conc: 23.39 ng/ml



#41 AR-1268-1


R.T.: 8.128 min
Delta R.T.: 0.000 min
Response: 11006162
Conc: 44.14 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT3-2022




#41 AR-1268-1

R.T.: 6.987 min
Delta R.T.: -0.001 min
Response: 12290813
Conc: 43.69 ng/ml




#42 AR-1268-2

R.T.: 8.218 min
Delta R.T.: -0.002 min
Response: 10017676
Conc: 43.88 ng/ml



#42 AR-1268-2


R.T.: 7.056 min
Delta R.T.: -0.002 min
Response: 10117257
Conc: 39.55 ng/ml



#43 AR-1268-3


R.T.: 8.435 min
Delta R.T.: -0.002 min
Response: 8567522
Conc: 44.49 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT3-2022




#43 AR-1268-3

R.T.: 7.275 min
Delta R.T.: -0.001 min
Response: 8670103
Conc: 40.29 ng/ml




#44 AR-1268-4

R.T.: 8.836 min
Delta R.T.: -0.002 min
Response: 3829657
Conc: 43.42 ng/ml



#44 AR-1268-4


R.T.: 7.589 min
Delta R.T.: -0.002 min
Response: 4071707
Conc: 42.85 ng/ml



#45 AR-1268-5

R.T.: 9.218 min
Delta R.T.: -0.002 min
Response: 25211285
Conc: 41.23 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOD-MDL-WATER-01-QT3-2022



#45 AR-1268-5

R.T.: 7.883 min
Delta R.T.: -0.002 min
Response: 25815049
Conc: 37.89 ng/ml