

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080622\
 Data File : PR055692.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Aug 2022 02:17
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 04:05:33 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:39:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.635	2.905	218.7E6	102.5E6	51.243	51.047
2) SA Decachlor...	9.526	8.120	84665821	86271361	47.617	47.426
Target Compounds						
3) L1 AR-1016-1	4.856	4.013	56371485	32435691	486.825	500.845
4) L1 AR-1016-2	4.878	4.030	80902329	45636879	492.960	499.665
5) L1 AR-1016-3	4.942	4.207	48889132	24733837	494.335	485.066
6) L1 AR-1016-4	5.044	4.258	39163359	18659997	494.404	491.154
7) L1 AR-1016-5	5.359	4.473	35494918	23772449	490.921	481.913
31) L7 AR-1260-1	6.571	5.556	51634466	49121273	476.942	487.520
32) L7 AR-1260-2	6.855	5.760	58849768	59281761	475.610	484.874
33) L7 AR-1260-3	7.245	5.917	43120365	54905278	477.698	485.480
34) L7 AR-1260-4	7.487	6.419	49069894	46177715	477.944	482.543
35) L7 AR-1260-5	7.825	6.684	92544980	108.2E6	476.082	481.746



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080622\
Data File : PR055692.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2022 02:17
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 04:05:33 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:39:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.635 min
Delta R.T.: 0.000 min
Response: 218656874
Conc: 51.24 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 2.905 min
Delta R.T.: 0.000 min
Response: 102493741
Conc: 51.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.526 min
Delta R.T.: -0.002 min
Response: 84665821
Conc: 47.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.120 min
Delta R.T.: -0.003 min
Response: 86271361
Conc: 47.43 ng/ml



#3 AR-1016-1


R.T.: 4.856 min
Delta R.T.: 0.000 min
Response: 56371485
Conc: 486.83 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.013 min
Delta R.T.: -0.001 min
Response: 32435691
Conc: 500.85 ng/ml




#4 AR-1016-2

R.T.: 4.878 min
Delta R.T.: 0.000 min
Response: 80902329
Conc: 492.96 ng/ml



#4 AR-1016-2


R.T.: 4.030 min
Delta R.T.: -0.001 min
Response: 45636879
Conc: 499.66 ng/ml



#5 AR-1016-3


R.T.: 4.942 min
Delta R.T.: -0.001 min
Response: 48889132
Conc: 494.34 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.207 min
Delta R.T.: -0.001 min
Response: 24733837
Conc: 485.07 ng/ml




#6 AR-1016-4

R.T.: 5.044 min
Delta R.T.: -0.001 min
Response: 39163359
Conc: 494.40 ng/ml



#6 AR-1016-4


R.T.: 4.258 min
Delta R.T.: -0.001 min
Response: 18659997
Conc: 491.15 ng/ml



#7 AR-1016-5


R.T.: 5.359 min
Delta R.T.: -0.002 min
Response: 35494918
Conc: 490.92 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 4.473 min
Delta R.T.: -0.002 min
Response: 23772449
Conc: 481.91 ng/ml




#31 AR-1260-1

R.T.: 6.571 min
Delta R.T.: -0.002 min
Response: 51634466
Conc: 476.94 ng/ml



#31 AR-1260-1


R.T.: 5.556 min
Delta R.T.: -0.001 min
Response: 49121273
Conc: 487.52 ng/ml



#32 AR-1260-2


R.T.: 6.855 min
Delta R.T.: 0.000 min
Response: 58849768
Conc: 475.61 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 5.760 min
Delta R.T.: -0.001 min
Response: 59281761
Conc: 484.87 ng/ml




#33 AR-1260-3

R.T.: 7.245 min
Delta R.T.: 0.000 min
Response: 43120365
Conc: 477.70 ng/ml



#33 AR-1260-3


R.T.: 5.917 min
Delta R.T.: -0.002 min
Response: 54905278
Conc: 485.48 ng/ml



#34 AR-1260-4


R.T.: 7.487 min
Delta R.T.: 0.000 min
Response: 49069894
Conc: 477.94 ng/ml

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.419 min
Delta R.T.: -0.001 min
Response: 46177715
Conc: 482.54 ng/ml



#35 AR-1260-5

R.T.: 7.825 min
Delta R.T.: -0.001 min
Response: 92544980
Conc: 476.08 ng/ml



#35 AR-1260-5

R.T.: 6.684 min
Delta R.T.: -0.002 min
Response: 108178306
Conc: 481.75 ng/ml