

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080822\
 Data File : PR055792.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2022 06:39
 Operator : AJ\MA
 Sample : N4081-01
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 2616

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/09/2022
 Supervised By :Ankita Jodhani 08/09/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 08:44:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:39:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.635	2.907	70022778	34327226	16.410	17.097
2) SA Decachlor...	9.547f	8.140f	26990656	28316209	15.180	15.566
Target Compounds						
26) L6 AR-1254-1	5.768	4.851	29192641	36283058	285.040	332.061
27) L6 AR-1254-2	6.007	5.013	44677318	24382575	303.925	251.805
28) L6 AR-1254-3	6.401	5.432f	34256299	29281620	233.819	192.234
29) L6 AR-1254-4	6.717	5.682f	17873133	15941635	172.406	161.787
30) L6 AR-1254-5	7.172	6.124f	23419831	24082636	215.989m	177.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080822\
Data File : PR055792.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2022 06:39
Operator : AJ\MA
Sample : N4081-01
Misc :
ALS Vial : 38 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
2616

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/09/2022
Supervised By :Ankita Jodhani 08/09/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 08:44:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:39:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

