

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081023\
 Data File : PR062560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2023 11:41
 Operator : YP\AJ
 Sample : 03572-08
 Misc : AR1262 LOQ 50 PPB
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOQ-WATER-02-QT3-2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 22:01:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 09 01:27:38 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.684	2.955	29894277	47507463	14.807	16.365
2) SA Decachlor...	9.618	8.235	27468944	71249476	18.442	17.171
Target Compounds						
36) L8 AR-1262-1	7.309	6.239	4852831	10736472	42.274	41.242
37) L8 AR-1262-2	7.890	6.780	7208181	18882027	39.889	38.846
38) L8 AR-1262-3	8.198	7.086	4674226	8182736	36.846	41.180
39) L8 AR-1262-4	8.284	7.154	3671632	14560902	36.383	39.114
40) L8 AR-1262-5	8.914	7.692	2442717	6686142	37.915	37.704



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081023\
Data File : PR062560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2023 11:41
Operator : YP\AJ
Sample : 03572-08
Misc : AR1262 LOQ 50 PPB
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT3-2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 22:01:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080923.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 09 01:27:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 29894277
Conc: 14.81 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT3-2023




#1 Tetrachloro-m-xylene

R.T.: 2.955 min
Delta R.T.: 0.000 min
Response: 47507463
Conc: 16.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.618 min
Delta R.T.: 0.000 min
Response: 27468944
Conc: 18.44 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.235 min
Delta R.T.: -0.002 min
Response: 71249476
Conc: 17.17 ng/ml



#36 AR-1262-1


R.T.: 7.309 min
Delta R.T.: -0.001 min
Response: 4852831
Conc: 42.27 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT3-2023




#36 AR-1262-1

R.T.: 6.239 min
Delta R.T.: -0.001 min
Response: 10736472
Conc: 41.24 ng/ml




#37 AR-1262-2

R.T.: 7.890 min
Delta R.T.: 0.000 min
Response: 7208181
Conc: 39.89 ng/ml



#37 AR-1262-2


R.T.: 6.780 min
Delta R.T.: -0.002 min
Response: 18882027
Conc: 38.85 ng/ml



#38 AR-1262-3


R.T.: 8.198 min
Delta R.T.: -0.001 min
Response: 4674226
Conc: 36.85 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT3-2023




#38 AR-1262-3

R.T.: 7.086 min
Delta R.T.: -0.001 min
Response: 8182736
Conc: 41.18 ng/ml




#39 AR-1262-4

R.T.: 8.284 min
Delta R.T.: -0.002 min
Response: 3671632
Conc: 36.38 ng/ml



#39 AR-1262-4


R.T.: 7.154 min
Delta R.T.: 0.000 min
Response: 14560902
Conc: 39.11 ng/ml



#40 AR-1262-5

R.T.: 8.914 min
Delta R.T.: 0.000 min
Response: 2442717
Conc: 37.91 ng/ml

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT3-2023



#40 AR-1262-5

R.T.: 7.692 min
Delta R.T.: -0.002 min
Response: 6686142
Conc: 37.70 ng/ml