

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081219\
 Data File : PR040272.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Aug 2019 10:19
 Operator : SM\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 12 12:20:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081019.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Aug 10 07:38:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.748	4.590	41406299	190.9E6	17.278	18.339
2) SA Decachlor...	8.675	10.310	34295993	156.2E6	14.094	17.836 #
Target Compounds						
9) L2 AR-1221-2	4.044	0.000	584273	0	28.832	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081219\
Data File : PR040272.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Aug 2019 10:19
Operator : SM\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 12 12:20:42 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081019.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Aug 10 07:38:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.748 min
Delta R.T.: 0.000 min
Response: 41406299
Conc: 17.28 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 4.590 min
Delta R.T.: 0.000 min
Response: 190924337
Conc: 18.34 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.675 min
Delta R.T.: -0.002 min
Response: 34295993
Conc: 14.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 10.310 min
Delta R.T.: -0.004 min
Response: 156199108
Conc: 17.84 ng/ml



#9 AR-1221-2

R.T.: 4.044 min
Delta R.T.: 0.004 min
Response: 584273
Conc: 28.83 ng/ml

Instrument :
ECD_R
ClientSampleId :
I.BLK



#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.881 min
Response: 0
Conc: N.D.