

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081519\
 Data File : PR040398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2019 02:20
 Operator : SM\AJ
 Sample : K3591-08
 Misc : AR1254 LOQ WATER 50PPB
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 LOQ-WATER-02-QT3-2019

Manual Integrations
 APPROVED

Ankita
 8/16/2019 2:52:52 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 05:27:11 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 15 15:16:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.755	4.599	17750725	86534636	14.446m	14.030
2)	SA Decachlor...	8.677	10.315	60599537	237.6E6	21.036	20.474
Target Compounds							
26)	L6 AR-1254-1	5.603	6.593	5476810	15171137	40.341	38.429
27)	L6 AR-1254-2	5.747	6.808	5146400	25085494	43.140	39.678
28)	L6 AR-1254-3	6.144	7.172	8728762	28043660	41.003	38.818
29)	L6 AR-1254-4	6.368	7.453	6874659	21924766	41.704	39.925
30)	L6 AR-1254-5	6.778	7.866	8730835	27246145	43.656	45.654

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081519\
Data File : PR040398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2019 02:20
Operator : SM\AJ
Sample : K3591-08
Misc : AR1254 LOQ WATER 50PPB
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
LOQ-WATER-02-QT3-2019

Manual Integrations
APPROVED

Ankita
8/16/2019 2:52:52 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 05:27:11 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081519.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 15 15:16:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

