

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081622\
 Data File : PR055899.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2022 15:05
 Operator : AJ\MA
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12543290

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 17 02:44:15 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081022CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 10 07:34:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.630	2.901	81889218	38099962	17.391	17.126
2) SA Decachlor...	9.521	8.115	57308662	59283861	34.417	34.012
Target Compounds						
26) L6 AR-1254-1	5.759	4.834	37550989	38961624	345.491	345.346
27) L6 AR-1254-2	5.995	4.992	53796753	33673949	338.650	345.818
28) L6 AR-1254-3	6.388	5.410	53837446	54529732	336.060	343.511
29) L6 AR-1254-4	6.700	5.655	39869956	34784443	339.718	340.886
30) L6 AR-1254-5	7.154	6.097	40761699	49870291	337.432	346.673

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081622\
Data File : PR055899.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2022 15:05
Operator : AJ\MA
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12543290

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 17 02:44:15 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081022CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 10 07:34:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

