

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081624\
 Data File : PR068337.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2024 10:29
 Operator : AJ\MA
 Sample : P3391-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-WATER-QT3-2024-06

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 02:34:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081424CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 14 10:38:14 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.675	3.008	36125504	92489467	15.174	17.015
2)	SA Decachlor...	9.584	8.323	99614973	114.5E6	39.724	40.615
Target Compounds							
3)	L1 AR-1016-1	4.899	4.138	1232825	4539113	23.434	25.297
4)	L1 AR-1016-2	4.920	4.157	2126717	6256602	21.622	24.856
5)	L1 AR-1016-3	4.984	4.337	1553756	3160296	24.167	23.064
6)	L1 AR-1016-4	5.087	4.389	1137542	2714638	21.349	24.274
7)	L1 AR-1016-5	5.402	4.608	845954	3483074	20.007	23.796
8)	L2 AR-1221-1	0.000	3.235	0	351724	N.D.	5.607 #
9)	L2 AR-1221-2	3.986	3.316	1077919	1834645	59.384	41.633 #
10)	L2 AR-1221-3	4.065	3.398	790307	2098629	12.999	14.657
11)	L3 AR-1232-1	4.065	3.398	790307	2098629	16.735	18.534
12)	L3 AR-1232-2	4.612	4.157	1340089	6256602	57.939	54.908
13)	L3 AR-1232-3	4.920	4.337	2126717	3160296	46.578	51.647
14)	L3 AR-1232-4	5.087	4.431	1137542	3364232	49.167	60.538
15)	L3 AR-1232-5	5.188	4.608	899250	3483074	82.997	58.200 #
16)	L4 AR-1242-1	4.899	4.138	1232825	4539113	27.813	29.731
17)	L4 AR-1242-2	4.920	4.157	2126717	6256602	25.647	29.463
18)	L4 AR-1242-3	4.984	4.337	1553756	3160296	28.776	27.295
19)	L4 AR-1242-4	5.087	4.431	1137542	3364232	24.956	29.201
20)	L4 AR-1242-5	5.876	4.983	462943	3612078	9.778	24.466 #
21)	L5 AR-1248-1	4.899	4.138	1232825	4539113	36.775	40.432
22)	L5 AR-1248-2	5.188	4.389	899250	2714638	22.120	16.307 #
23)	L5 AR-1248-3	5.402	4.431	845954	3364232	14.404	19.846 #
24)	L5 AR-1248-4	5.836	4.608	299333	3483074	4.392	17.101 #
25)	L5 AR-1248-5	5.876	5.025	462943	897709	5.893	4.490
26)	L6 AR-1254-1	5.806	4.983	1406901	3612078	23.423	11.783 #
27)	L6 AR-1254-2	6.042	5.143	1101094	2996713	11.038	10.992
28)	L6 AR-1254-3	6.433	5.568	703866	1189534	5.447	2.747 #
29)	L6 AR-1254-4	6.749	5.817	714602	472465	6.397	1.776 #
30)	L6 AR-1254-5	7.199	6.266	4713312	9793115	37.550	24.870 #
31)	L7 AR-1260-1	6.616	5.713	1994089	7892775	24.825	25.917
32)	L7 AR-1260-2	6.898	5.918	3986312	9483521	27.372	25.682
33)	L7 AR-1260-3	7.287	6.078	4564499	8772739	27.704	25.700
34)	L7 AR-1260-4	7.529	6.589	3617790	8061531	26.764	27.350
35)	L7 AR-1260-5	7.868	6.855	11804173	16629860	25.181	24.697
36)	L8 AR-1262-1	7.287	6.313	4564499	8339440	19.507	20.285
37)	L8 AR-1262-2	7.868	6.855	11804173	16629860	22.788	22.821
38)	L8 AR-1262-3	8.180	7.164	6459274	4557015	18.186	15.535
39)	L8 AR-1262-4	8.261	7.229	4635783	11951544	15.504	22.006 #
40)	L8 AR-1262-5	8.886	7.771	3404858	4319606	15.809	17.231
41)	L9 AR-1268-1	8.180	7.164	6459274	4557015	8.403	5.062 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081624\
Data File : PR068337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2024 10:29
Operator : AJ\MA
Sample : P3391-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 02:34:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081424CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 14 10:38:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.261	7.229	4635783	11951544	6.710	14.499 #
43)	L9 AR-1268-3	8.490	7.460	615112	325937	0.945	0.468 #
44)	L9 AR-1268-4	8.886	7.771	3404858	4319606	13.407	14.399
45)	L9 AR-1268-5	9.272	8.071	1628797	2075909	0.872	0.993



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081624\
Data File : PR068337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2024 10:29
Operator : AJ\MA
Sample : P3391-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 02:34:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081424CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 14 10:38:14 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 AR-1016-1


R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 1232825
Conc: 23.43 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#3 AR-1016-1

R.T.: 4.138 min
Delta R.T.: -0.002 min
Response: 4539113
Conc: 25.30 ng/ml




#4 AR-1016-2

R.T.: 4.920 min
Delta R.T.: -0.002 min
Response: 2126717
Conc: 21.62 ng/ml



#4 AR-1016-2


R.T.: 4.157 min
Delta R.T.: -0.002 min
Response: 6256602
Conc: 24.86 ng/ml



#5 AR-1016-3


R.T.: 4.984 min
Delta R.T.: -0.002 min
Response: 1553756
Conc: 24.17 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#5 AR-1016-3

R.T.: 4.337 min
Delta R.T.: -0.002 min
Response: 3160296
Conc: 23.06 ng/ml




#6 AR-1016-4

R.T.: 5.087 min
Delta R.T.: 0.000 min
Response: 1137542
Conc: 21.35 ng/ml



#6 AR-1016-4

R.T.: 4.389 min
Delta R.T.: -0.002 min
Response: 2714638
Conc: 24.27 ng/ml



#7 AR-1016-5

R.T.: 5.402 min
Delta R.T.: -0.002 min
Response: 845954
Conc: 20.01 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06

#7 AR-1016-5


R.T.: 4.608 min
Delta R.T.: -0.002 min
Response: 3483074
Conc: 23.80 ng/ml

#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 3.899 min
Response: 0
Conc: N.D.

#8 AR-1221-1


R.T.: 3.235 min
Delta R.T.: 0.000 min
Response: 351724
Conc: 5.61 ng/ml



#9 AR-1221-2


R.T.: 3.986 min
Delta R.T.: 0.000 min
Response: 1077919
Conc: 59.38 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#9 AR-1221-2

R.T.: 3.316 min
Delta R.T.: -0.006 min
Response: 1834645
Conc: 41.63 ng/ml




#10 AR-1221-3

R.T.: 4.065 min
Delta R.T.: 0.000 min
Response: 790307
Conc: 13.00 ng/ml



#10 AR-1221-3


R.T.: 3.398 min
Delta R.T.: -0.001 min
Response: 2098629
Conc: 14.66 ng/ml



#11 AR-1232-1


R.T.: 4.065 min
Delta R.T.: 0.000 min
Response: 790307
Conc: 16.74 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#11 AR-1232-1

R.T.: 3.398 min
Delta R.T.: -0.001 min
Response: 2098629
Conc: 18.53 ng/ml




#12 AR-1232-2

R.T.: 4.612 min
Delta R.T.: -0.002 min
Response: 1340089
Conc: 57.94 ng/ml



#12 AR-1232-2


R.T.: 4.157 min
Delta R.T.: -0.002 min
Response: 6256602
Conc: 54.91 ng/ml



#13 AR-1232-3


R.T.: 4.920 min
Delta R.T.: -0.002 min
Response: 2126717
Conc: 46.58 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#13 AR-1232-3

R.T.: 4.337 min
Delta R.T.: -0.002 min
Response: 3160296
Conc: 51.65 ng/ml




#14 AR-1232-4

R.T.: 5.087 min
Delta R.T.: -0.002 min
Response: 1137542
Conc: 49.17 ng/ml



#14 AR-1232-4


R.T.: 4.431 min
Delta R.T.: -0.002 min
Response: 3364232
Conc: 60.54 ng/ml



#15 AR-1232-5


R.T.: 5.188 min
Delta R.T.: -0.004 min
Response: 899250
Conc: 83.00 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#15 AR-1232-5

R.T.: 4.608 min
Delta R.T.: -0.002 min
Response: 3483074
Conc: 58.20 ng/ml




#16 AR-1242-1

R.T.: 4.899 min
Delta R.T.: -0.001 min
Response: 1232825
Conc: 27.81 ng/ml



#16 AR-1242-1


R.T.: 4.138 min
Delta R.T.: -0.002 min
Response: 4539113
Conc: 29.73 ng/ml



#17 AR-1242-2


R.T.: 4.920 min
Delta R.T.: -0.002 min
Response: 2126717
Conc: 25.65 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#17 AR-1242-2

R.T.: 4.157 min
Delta R.T.: -0.002 min
Response: 6256602
Conc: 29.46 ng/ml




#18 AR-1242-3

R.T.: 4.984 min
Delta R.T.: -0.002 min
Response: 1553756
Conc: 28.78 ng/ml



#18 AR-1242-3


R.T.: 4.337 min
Delta R.T.: -0.002 min
Response: 3160296
Conc: 27.29 ng/ml



#19 AR-1242-4


R.T.: 5.087 min
Delta R.T.: -0.002 min
Response: 1137542
Conc: 24.96 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#19 AR-1242-4

R.T.: 4.431 min
Delta R.T.: -0.002 min
Response: 3364232
Conc: 29.20 ng/ml




#20 AR-1242-5

R.T.: 5.876 min
Delta R.T.: 0.000 min
Response: 462943
Conc: 9.78 ng/ml



#20 AR-1242-5


R.T.: 4.983 min
Delta R.T.: -0.002 min
Response: 3612078
Conc: 24.47 ng/ml



#21 AR-1248-1


R.T.: 4.899 min
Delta R.T.: -0.001 min
Response: 1232825
Conc: 36.78 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#21 AR-1248-1

R.T.: 4.138 min
Delta R.T.: -0.002 min
Response: 4539113
Conc: 40.43 ng/ml




#22 AR-1248-2

R.T.: 5.188 min
Delta R.T.: -0.003 min
Response: 899250
Conc: 22.12 ng/ml



#22 AR-1248-2


R.T.: 4.389 min
Delta R.T.: -0.001 min
Response: 2714638
Conc: 16.31 ng/ml



#23 AR-1248-3


R.T.: 5.402 min
Delta R.T.: -0.002 min
Response: 845954
Conc: 14.40 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#23 AR-1248-3

R.T.: 4.431 min
Delta R.T.: -0.002 min
Response: 3364232
Conc: 19.85 ng/ml




#24 AR-1248-4

R.T.: 5.836 min
Delta R.T.: 0.000 min
Response: 299333
Conc: 4.39 ng/ml



#24 AR-1248-4


R.T.: 4.608 min
Delta R.T.: -0.002 min
Response: 3483074
Conc: 17.10 ng/ml



#25 AR-1248-5


R.T.: 5.876 min
Delta R.T.: 0.000 min
Response: 462943
Conc: 5.89 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#25 AR-1248-5

R.T.: 5.025 min
Delta R.T.: -0.002 min
Response: 897709
Conc: 4.49 ng/ml




#26 AR-1254-1

R.T.: 5.806 min
Delta R.T.: -0.001 min
Response: 1406901
Conc: 23.42 ng/ml



#26 AR-1254-1


R.T.: 4.983 min
Delta R.T.: -0.002 min
Response: 3612078
Conc: 11.78 ng/ml



#27 AR-1254-2


R.T.: 6.042 min
Delta R.T.: -0.002 min
Response: 1101094
Conc: 11.04 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#27 AR-1254-2

R.T.: 5.143 min
Delta R.T.: -0.003 min
Response: 2996713
Conc: 10.99 ng/ml




#28 AR-1254-3

R.T.: 6.433 min
Delta R.T.: -0.002 min
Response: 703866
Conc: 5.45 ng/ml



#28 AR-1254-3


R.T.: 5.568 min
Delta R.T.: -0.004 min
Response: 1189534
Conc: 2.75 ng/ml



#29 AR-1254-4


R.T.: 6.749 min
Delta R.T.: 0.003 min
Response: 714602
Conc: 6.40 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#29 AR-1254-4

R.T.: 5.817 min
Delta R.T.: -0.003 min
Response: 472465
Conc: 1.78 ng/ml




#30 AR-1254-5

R.T.: 7.199 min
Delta R.T.: -0.001 min
Response: 4713312
Conc: 37.55 ng/ml



#30 AR-1254-5


R.T.: 6.266 min
Delta R.T.: -0.003 min
Response: 9793115
Conc: 24.87 ng/ml



#31 AR-1260-1


R.T.: 6.616 min
Delta R.T.: 0.000 min
Response: 1994089
Conc: 24.83 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#31 AR-1260-1

R.T.: 5.713 min
Delta R.T.: -0.002 min
Response: 7892775
Conc: 25.92 ng/ml




#32 AR-1260-2

R.T.: 6.898 min
Delta R.T.: 0.000 min
Response: 3986312
Conc: 27.37 ng/ml



#32 AR-1260-2


R.T.: 5.918 min
Delta R.T.: -0.003 min
Response: 9483521
Conc: 25.68 ng/ml



#33 AR-1260-3


R.T.: 7.287 min
Delta R.T.: 0.000 min
Response: 4564499
Conc: 27.70 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#33 AR-1260-3

R.T.: 6.078 min
Delta R.T.: -0.002 min
Response: 8772739
Conc: 25.70 ng/ml




#34 AR-1260-4

R.T.: 7.529 min
Delta R.T.: 0.000 min
Response: 3617790
Conc: 26.76 ng/ml



#34 AR-1260-4


R.T.: 6.589 min
Delta R.T.: -0.002 min
Response: 8061531
Conc: 27.35 ng/ml



#35 AR-1260-5


R.T.: 7.868 min
Delta R.T.: 0.000 min
Response: 11804173
Conc: 25.18 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#35 AR-1260-5

R.T.: 6.855 min
Delta R.T.: -0.003 min
Response: 16629860
Conc: 24.70 ng/ml




#36 AR-1262-1

R.T.: 7.287 min
Delta R.T.: 0.000 min
Response: 4564499
Conc: 19.51 ng/ml



#36 AR-1262-1


R.T.: 6.313 min
Delta R.T.: -0.002 min
Response: 8339440
Conc: 20.29 ng/ml



#37 AR-1262-2


R.T.: 7.868 min
Delta R.T.: 0.002 min
Response: 11804173
Conc: 22.79 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#37 AR-1262-2

R.T.: 6.855 min
Delta R.T.: 0.000 min
Response: 16629860
Conc: 22.82 ng/ml




#38 AR-1262-3

R.T.: 8.180 min
Delta R.T.: 0.009 min
Response: 6459274
Conc: 18.19 ng/ml



#38 AR-1262-3


R.T.: 7.164 min
Delta R.T.: 0.000 min
Response: 4557015
Conc: 15.54 ng/ml



#39 AR-1262-4


R.T.: 8.261 min
Delta R.T.: 0.003 min
Response: 4635783
Conc: 15.50 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#39 AR-1262-4

R.T.: 7.229 min
Delta R.T.: -0.002 min
Response: 11951544
Conc: 22.01 ng/ml




#40 AR-1262-5

R.T.: 8.886 min
Delta R.T.: 0.003 min
Response: 3404858
Conc: 15.81 ng/ml



#40 AR-1262-5


R.T.: 7.771 min
Delta R.T.: 0.000 min
Response: 4319606
Conc: 17.23 ng/ml



#41 AR-1268-1


R.T.: 8.180 min
Delta R.T.: 0.008 min
Response: 6459274
Conc: 8.40 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#41 AR-1268-1

R.T.: 7.164 min
Delta R.T.: 0.000 min
Response: 4557015
Conc: 5.06 ng/ml




#42 AR-1268-2

R.T.: 8.261 min
Delta R.T.: -0.001 min
Response: 4635783
Conc: 6.71 ng/ml



#42 AR-1268-2


R.T.: 7.229 min
Delta R.T.: -0.004 min
Response: 11951544
Conc: 14.50 ng/ml



#43 AR-1268-3


R.T.: 8.490 min
Delta R.T.: 0.011 min
Response: 615112
Conc: 0.95 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06




#43 AR-1268-3

R.T.: 7.460 min
Delta R.T.: 0.002 min
Response: 325937
Conc: 0.47 ng/ml




#44 AR-1268-4

R.T.: 8.886 min
Delta R.T.: 0.001 min
Response: 3404858
Conc: 13.41 ng/ml



#44 AR-1268-4


R.T.: 7.771 min
Delta R.T.: 0.000 min
Response: 4319606
Conc: 14.40 ng/ml



#45 AR-1268-5

R.T.: 9.272 min
Delta R.T.: 0.002 min
Response: 1628797
Conc: 0.87 ng/ml

Instrument :
ECD_R
ClientSampleId :
MDL-WATER-QT3-2024-06



#45 AR-1268-5

R.T.: 8.071 min
Delta R.T.: 0.002 min
Response: 2075909
Conc: 0.99 ng/ml