

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081822\
 Data File : PR055971.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2022 15:26
 Operator : AJ\MA
 Sample : N4210-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 BGC48

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/19/2022
 Supervised By :Ankita Jodhani 08/19/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 22:47:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081022CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 10 07:34:29 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.629	2.900	81453193	41253925	17.298	18.544
2) SA Decachlor...	9.522	8.115	59304750	52874779	35.616	30.335
Target Compounds						
31) L7 AR-1260-1	6.567	5.552	18735951	12348021	165.091	115.232 #
32) L7 AR-1260-2	6.849	5.751	8520617	8949429	65.407	69.158
33) L7 AR-1260-3	7.240	5.908	10691845	6900525	112.566	56.258 #
34) L7 AR-1260-4	7.483	6.412	11125437	8647236	102.212m	84.026
35) L7 AR-1260-5	7.822	6.676	18492780	16509544	89.681	68.582

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081822\
Data File : PR055971.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2022 15:26
Operator : AJ\MA
Sample : N4210-05
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
BGC48


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/19/2022
Supervised By :Ankita Jodhani 08/19/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 22:47:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR081022CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 10 07:34:29 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#31 AR-1260-1

R.T.: 6.567 min
Delta R.T.: -0.002 min
Response: 18735951
Conc: 165.09 ng/ml

Instrument :
ECD_R
ClientSampleId :
BGC48


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/19/2022
Supervised By :Ankita Jodhani 08/19/2022




#31 AR-1260-1

R.T.: 5.552 min
Delta R.T.: 0.000 min
Response: 12348021
Conc: 115.23 ng/ml




#32 AR-1260-2

R.T.: 6.849 min
Delta R.T.: -0.003 min
Response: 8520617
Conc: 65.41 ng/ml



#32 AR-1260-2

R.T.: 5.751 min
Delta R.T.: -0.005 min
Response: 8949429
Conc: 69.16 ng/ml




#33 AR-1260-3

R.T.: 7.240 min
Delta R.T.: -0.002 min
Response: 10691845
Conc: 112.57 ng/ml

Instrument :
ECD_R
ClientSampleId :
BGC48


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/19/2022
Supervised By :Ankita Jodhani 08/19/2022




#33 AR-1260-3

R.T.: 5.908 min
Delta R.T.: -0.005 min
Response: 6900525
Conc: 56.26 ng/ml




#34 AR-1260-4

R.T.: 7.483 min
Delta R.T.: -0.003 min
Response: 11125437
Conc: 102.21 ng/ml m



#34 AR-1260-4

R.T.: 6.412 min
Delta R.T.: -0.004 min
Response: 8647236
Conc: 84.03 ng/ml




#35 AR-1260-5

R.T.: 7.822 min
Delta R.T.: -0.002 min
Response: 18492780
Conc: 89.68 ng/ml

Instrument :
ECD_R
ClientSampleId :
BGC48

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/19/2022
Supervised By :Ankita Jodhani 08/19/2022



#35 AR-1260-5

R.T.: 6.676 min
Delta R.T.: -0.005 min
Response: 16509544
Conc: 68.58 ng/ml