

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR081920CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 20 08:59:01 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046782.D 1600 =PR046784.D 800 =PR046783.D
 200 =PR046781.D 100 =PR046780.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.285	2.093	2.273	2.465	2.130	2.249	E6	6.56
2)	SA	Decachlorobiphenyl	2.104	1.899	2.080	2.262	2.296	2.128	E6	7.47
3)	L1	AR-1016-1	9.063	7.973	8.885	9.869	8.576	8.873	E4	7.82
4)	L1	AR-1016-2	1.268	1.140	1.240	1.363	1.190	1.240	E5	6.81
5)	L1	AR-1016-3	7.837	6.748	7.554	8.554	7.376	7.614	E4	8.67
6)	L1	AR-1016-4	6.572	5.741	6.374	7.127	6.205	6.404	E4	7.93
7)	L1	AR-1016-5	6.728	5.706	6.409	7.299	6.279	6.484	E4	9.05
8)	L2	AR-1221-1					2.759	2.759	E4	0.00
9)	L2	AR-1221-2					2.084	2.084	E4	0.00
10)	L2	AR-1221-3					6.692	6.692	E4	0.00
11)	L3	AR-1232-1					5.787	5.787	E4	0.00
12)	L3	AR-1232-2					3.009	3.009	E4	0.00
13)	L3	AR-1232-3					6.146	6.146	E4	0.00
14)	L3	AR-1232-4					3.095	3.095	E4	0.00
15)	L3	AR-1232-5					2.512	2.512	E4	0.00
16)	L4	AR-1242-1					7.913	7.913	E4	0.00
17)	L4	AR-1242-2					1.086	1.086	E5	0.00
18)	L4	AR-1242-3					6.764	6.764	E4	0.00
19)	L4	AR-1242-4					5.366	5.366	E4	0.00
20)	L4	AR-1242-5					6.637	6.637	E4	0.00
21)	L5	AR-1248-1					5.096	5.096	E4	0.00
22)	L5	AR-1248-2					7.375	7.375	E4	0.00
23)	L5	AR-1248-3					8.189	8.189	E4	0.00
24)	L5	AR-1248-4					1.017	1.017	E5	0.00
25)	L5	AR-1248-5					9.648	9.648	E4	0.00
26)	L6	AR-1254-1					1.168	1.168	E5	0.00
27)	L6	AR-1254-2					1.900	1.900	E5	0.00
28)	L6	AR-1254-3					2.122	2.122	E5	0.00
29)	L6	AR-1254-4					1.673	1.673	E5	0.00
30)	L6	AR-1254-5					1.650	1.650	E5	0.00
31)	L7	AR-1260-1	1.314	1.127	1.277	1.465	1.379	1.313	E5	9.58
32)	L7	AR-1260-2	1.572	1.364	1.541	1.713	1.596	1.557	E5	8.08
33)	L7	AR-1260-3	1.184	1.016	1.159	1.273	1.211	1.169	E5	8.16
34)	L7	AR-1260-4	1.392	1.200	1.353	1.499	1.412	1.371	E5	8.00
35)	L7	AR-1260-5	2.840	2.571	2.851	3.042	2.784	2.818	E5	5.99
36)	L8	AR-1262-1					7.465	7.465	E4	0.00
37)	L8	AR-1262-2					3.307	3.307	E5	0.00
38)	L8	AR-1262-3					1.452	1.452	E5	0.00
39)	L8	AR-1262-4					1.668	1.668	E5	0.00
40)	L8	AR-1262-5					1.198	1.198	E5	0.00
41)	L9	AR-1268-1					4.212	4.212	E5	0.00
42)	L9	AR-1268-2					3.802	3.802	E5	0.00
43)	L9	AR-1268-3					3.166	3.166	E5	0.00
44)	L9	AR-1268-4					1.545	1.545	E5	0.00
45)	L9	AR-1268-5					9.831	9.831	E5	0.00

Signal #2 Calibration Files

400 =PR046782.D 1600 =PR046784.D 800 =PR046783.D
 200 =PR046781.D 100 =PR046780.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	6.631	6.139	6.936	7.262	6.252	6.644	E6	7.05
2)	SA	Decachlorobiphenyl	1.409	1.315	1.392	1.470	1.426	1.402	E7	4.06
3)	L1	AR-1016-1	2.027	1.766	2.017	2.215	1.958	1.997	E5	8.06
4)	L1	AR-1016-2	3.397	2.986	3.382	3.707	3.308	3.356	E5	7.66
5)	L1	AR-1016-3	1.765	1.621	1.832	1.829	1.492	1.708	E5	8.67

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR081920CLP.M
 Title : GC EXTRACTABLES
 Last Update : Thu Aug 20 08:59:01 2020
 Response Via : Initial Calibration

Calibration Files

400 =PR046782.D 1600 =PR046784.D 800 =PR046783.D
 200 =PR046781.D 100 =PR046780.D

Compound			400	1600	800	200	100	Avg		%RSD
6)	L1	AR-1016-4	1.034	0.890	1.064	1.154	1.135	1.056	E5	9.92
7)	L1	AR-1016-5	1.633	1.430	1.641	1.798	1.695	1.639	E5	8.20
8)	L2	AR-1221-1					8.333	8.333	E4	0.00
9)	L2	AR-1221-2					6.589	6.589	E4	0.00
10)	L2	AR-1221-3					1.971	1.971	E5	0.00
11)	L3	AR-1232-1					1.434	1.434	E5	0.00
12)	L3	AR-1232-2					1.626	1.626	E5	0.00
13)	L3	AR-1232-3					6.532	6.532	E4	0.00
14)	L3	AR-1232-4					7.773	7.773	E4	0.00
15)	L3	AR-1232-5					8.464	8.464	E4	0.00
16)	L4	AR-1242-1					1.660	1.660	E5	0.00
17)	L4	AR-1242-2					2.747	2.747	E5	0.00
18)	L4	AR-1242-3					1.226	1.226	E5	0.00
19)	L4	AR-1242-4					1.139	1.139	E5	0.00
20)	L4	AR-1242-5					1.477	1.477	E5	0.00
21)	L5	AR-1248-1					1.098	1.098	E5	0.00
22)	L5	AR-1248-2					1.317	1.317	E5	0.00
23)	L5	AR-1248-3					1.527	1.527	E5	0.00
24)	L5	AR-1248-4					2.086	2.086	E5	0.00
25)	L5	AR-1248-5					2.464	2.464	E5	0.00
26)	L6	AR-1254-1					3.449	3.449	E5	0.00
27)	L6	AR-1254-2					3.303	3.303	E5	0.00
28)	L6	AR-1254-3					6.985	6.985	E5	0.00
29)	L6	AR-1254-4					5.018	5.018	E5	0.00
30)	L6	AR-1254-5					6.074	6.074	E5	0.00
31)	L7	AR-1260-1	3.249	2.817	3.237	3.622	3.169	3.219	E5	8.89
32)	L7	AR-1260-2	4.560	4.304	5.559	4.920	4.607	4.790	E5	10.07
33)	L7	AR-1260-3	4.135	3.878	4.552	4.369	4.002	4.187	E5	6.53
34)	L7	AR-1260-4	4.169	3.887	4.675	4.377	4.233	4.268	E5	6.76
35)	L7	AR-1260-5	1.590	1.634	1.962	1.657	1.542	1.677	E6	9.85
36)	L8	AR-1262-1					1.628	1.628	E5	0.00
37)	L8	AR-1262-2					1.705	1.705	E6	0.00
38)	L8	AR-1262-3					7.735	7.735	E5	0.00
39)	L8	AR-1262-4					1.347	1.347	E6	0.00
40)	L8	AR-1262-5					7.853	7.853	E5	0.00
41)	L9	AR-1268-1					2.498	2.498	E6	0.00
42)	L9	AR-1268-2					2.460	2.460	E6	0.00
43)	L9	AR-1268-3					2.176	2.176	E6	0.00
44)	L9	AR-1268-4					9.459	9.459	E5	0.00
45)	L9	AR-1268-5					6.828	6.828	E6	0.00

 (#) = Out of Range