

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081922\
 Data File : PR056013.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Aug 2022 19:54
 Operator : AJ\MA
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/22/2022
 Supervised By :Ankita Jodhani 08/23/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 20 04:19:03 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:39:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.630	2.900	209.1E6	97049726	48.995	48.336
2) SA Decachlor...	9.519	8.111	75976907	81390561	42.731	44.743m
Target Compounds						
3) L1 AR-1016-1	4.850	4.006	53632655	29976159	463.173	462.867
4) L1 AR-1016-2	4.872	4.024	76362392	42077928	465.297	460.699
5) L1 AR-1016-3	4.936	4.200	46525392	23872406	470.435	468.172
6) L1 AR-1016-4	5.038	4.252	36887467	17949736	465.673	472.459
7) L1 AR-1016-5	5.354	4.466	33669344	22933469	465.672	464.905
31) L7 AR-1260-1	6.566	5.548	48083753	47168438	444.144	468.139
32) L7 AR-1260-2	6.849	5.752	55099036	56852116	445.298	465.002
33) L7 AR-1260-3	7.239	5.909	40101185	52910307	444.251	467.840
34) L7 AR-1260-4	7.481	6.411	46335155	44519162	451.308	465.211
35) L7 AR-1260-5	7.820	6.676	86400728	103.9E6	444.474	462.496

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR081922\
Data File : PR056013.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Aug 2022 19:54
Operator : AJ\MA
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/22/2022
Supervised By :Ankita Jodhani 08/23/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 20 04:19:03 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:39:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

