

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082318\
 Data File : PR031577.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2018 16:06
 Operator : SM\SJ
 Sample : J4556-08RE
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 20180533-AMENDED-COMPOSITE-(

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 09:34:57 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082318.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 24 09:16:01 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.550	3.673	364.7E6	703.0E6	16.675	18.326
2) SA Decachlor...	10.325	8.550	317.9E6	217.5E6	10.243	10.701
Target Compounds						
26) L6 AR-1254-1	6.780	5.650	313.0E6	549.6E6	162.446	160.965
27) L6 AR-1254-2	7.060	5.958	169.2E6	349.2E6	138.641	143.401
28) L6 AR-1254-3	7.146	6.271	343.5E6	360.4E6	156.902	84.243 #
29) L6 AR-1254-4	7.429	6.579	189.6E6	171.6E6	104.072	86.118
30) L6 AR-1254-5	7.698	6.678	123.1E6	613.7E6	93.324	218.866 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082318\
Data File : PR031577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2018 16:06
Operator : SM\SJ
Sample : J4556-08RE
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
20180533-AMENDED-COMPOSITE-(

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 09:34:57 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082318.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 24 09:16:01 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

