

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR082318.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 24 09:16:01 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031556.D 1000 =PR031554.D 750 =PR031555.D
 250 =PR031557.D 50 =PR031558.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.300	2.196	2.187	2.219	2.032	2.187	E7	4.44
2) SA	Decachlorobiphenyl	3.078	2.827	2.766	3.184	3.663	3.104	E7	11.51
3) L1	AR-1016-1	5.901	4.822	5.513	5.897	5.616	5.550	E5	7.95
4) L1	AR-1016-2	6.118	5.534	5.662	6.028	5.676	5.804	E5	4.38
5) L1	AR-1016-3	9.545	9.130	8.805	9.466	9.897	9.369	E5	4.45
6) L1	AR-1016-4	8.320	7.739	7.628	8.253	8.297	8.047	E5	4.17
7) L1	AR-1016-5	7.245	6.766	6.544	7.138	7.635	7.066	E5	6.02
8) L2	AR-1221-1	2.591					2.591	E5	0.00
9) L2	AR-1221-2	2.108					2.108	E5	0.00
10) L2	AR-1221-3	6.020					6.020	E5	0.00
11) L3	AR-1232-1	6.081					6.081	E5	0.00
12) L3	AR-1232-2	3.297					3.297	E5	0.00
13) L3	AR-1232-3	2.368					2.368	E5	0.00
14) L3	AR-1232-4	2.357					2.357	E5	0.00
15) L3	AR-1232-5	3.471					3.471	E5	0.00
16) L4	AR-1242-1	2.450					2.450	E5	0.00
17) L4	AR-1242-2	4.788					4.788	E5	0.00
18) L4	AR-1242-3	2.479					2.479	E5	0.00
19) L4	AR-1242-4	1.873					1.873	E5	0.00
20) L4	AR-1242-5	6.247					6.247	E5	0.00
21) L5	AR-1248-1	7.980					7.980	E5	0.00
22) L5	AR-1248-2	8.381					8.381	E5	0.00
23) L5	AR-1248-3	1.129					1.129	E6	0.00
24) L5	AR-1248-4	1.064					1.064	E6	0.00
25) L5	AR-1248-5	7.096					7.096	E5	0.00
26) L6	AR-1254-1	1.927					1.927	E6	0.00
27) L6	AR-1254-2	1.220					1.220	E6	0.00
28) L6	AR-1254-3	2.189					2.189	E6	0.00
29) L6	AR-1254-4	1.822					1.822	E6	0.00
30) L6	AR-1254-5	1.319					1.319	E6	0.00
31) L7	AR-1260-1	1.581	1.458	1.432	1.585	1.684	1.548	E6	6.68
32) L7	AR-1260-2	2.123	1.941	1.916	2.119	2.230	2.066	E6	6.47
33) L7	AR-1260-3	2.281	1.915	2.124	2.281	2.229	2.166	E6	7.13
34) L7	AR-1260-4	1.640	1.482	1.477	1.689	1.803	1.618	E6	8.63
35) L7	AR-1260-5	3.726	3.550	3.404	3.630	3.661	3.594	E6	3.44
36) L8	AR-1262-1	1.077					1.077	E6	0.00
37) L8	AR-1262-2	1.033					1.033	E6	0.00
38) L8	AR-1262-3	1.102					1.102	E6	0.00
39) L8	AR-1262-4	2.250					2.250	E6	0.00
40) L8	AR-1262-5	1.601					1.601	E6	0.00
41) L9	AR-1268-1	7.406					7.406	E5	0.00
42) L9	AR-1268-2	1.116					1.116	E6	0.00
43) L9	AR-1268-3	4.446					4.446	E6	0.00
44) L9	AR-1268-4	3.123					3.123	E6	0.00
45) L9	AR-1268-5	9.270					9.270	E6	0.00

Signal #2 Calibration Files

500 =PR031556.D 1000 =PR031554.D 750 =PR031555.D
 250 =PR031557.D 50 =PR031558.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.045	3.926	3.840	3.914	3.458	3.836	E7	5.84
2) SA	Decachlorobiphenyl	1.980	1.778	1.791	2.079	2.533	2.032	E7	15.15
3) L1	AR-1016-1	1.173	1.078	1.052	1.197	1.272	1.154	E6	7.77
4) L1	AR-1016-2	1.776	1.617	1.548	1.661	1.871	1.694	E6	7.61
5) L1	AR-1016-3	2.971	2.912	2.739	3.075	3.268	2.993	E6	6.55

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\
 Method File : PR082318.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 24 09:16:01 2018
 Response Via : Initial Calibration

Calibration Files

500 =PR031556.D 1000 =PR031554.D 750 =PR031555.D
 250 =PR031557.D 50 =PR031558.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.843	1.737	1.665	1.799	1.806	1.770	E6	3.95
7)	L1 AR-1016-5	1.648	1.548	1.491	1.645	1.701	1.607	E6	5.29
8)	L2 AR-1221-1	4.757					4.757	E5	0.00
9)	L2 AR-1221-2	3.601					3.601	E5	0.00
10)	L2 AR-1221-3	1.148					1.148	E6	0.00
11)	L3 AR-1232-1	6.362					6.362	E5	0.00
12)	L3 AR-1232-2	1.387					1.387	E6	0.00
13)	L3 AR-1232-3	5.944					5.944	E5	0.00
14)	L3 AR-1232-4	7.736					7.736	E5	0.00
15)	L3 AR-1232-5	7.584					7.584	E5	0.00
16)	L4 AR-1242-1	5.778					5.778	E5	0.00
17)	L4 AR-1242-2	1.141					1.141	E6	0.00
18)	L4 AR-1242-3	9.259					9.259	E5	0.00
19)	L4 AR-1242-4	1.250					1.250	E6	0.00
20)	L4 AR-1242-5	8.354					8.354	E5	0.00
21)	L5 AR-1248-1	1.622					1.622	E6	0.00
22)	L5 AR-1248-2	3.243					3.243	E6	0.00
23)	L5 AR-1248-3	2.322					2.322	E6	0.00
24)	L5 AR-1248-4	2.331					2.331	E6	0.00
25)	L5 AR-1248-5	1.520					1.520	E6	0.00
26)	L6 AR-1254-1	3.415					3.415	E6	0.00
27)	L6 AR-1254-2	2.435					2.435	E6	0.00
28)	L6 AR-1254-3	4.278					4.278	E6	0.00
29)	L6 AR-1254-4	1.993					1.993	E6	0.00
30)	L6 AR-1254-5	2.804					2.804	E6	0.00
31)	L7 AR-1260-1	3.273	3.003	2.960	3.303	3.708	3.250	E6	9.20
32)	L7 AR-1260-2	3.948	3.596	3.542	4.054	4.147	3.857	E6	7.08
33)	L7 AR-1260-3	3.228	2.922	2.893	3.283	3.822	3.230	E6	11.60
34)	L7 AR-1260-4	1.912	1.666	1.671	2.012	2.366	1.925	E6	14.99
35)	L7 AR-1260-5	3.821	3.362	3.387	3.929	4.999	3.900	E6	17.04
36)	L8 AR-1262-1	2.264					2.264	E6	0.00
37)	L8 AR-1262-2	1.596					1.596	E6	0.00
38)	L8 AR-1262-3	8.425					8.425	E5	0.00
39)	L8 AR-1262-4	1.535					1.535	E6	0.00
40)	L8 AR-1262-5	7.929					7.929	E5	0.00
41)	L9 AR-1268-1	1.174					1.174	E6	0.00
42)	L9 AR-1268-2	2.843					2.843	E6	0.00
43)	L9 AR-1268-3	2.428					2.428	E6	0.00
44)	L9 AR-1268-4	6.812					6.812	E5	0.00
45)	L9 AR-1268-5	6.254					6.254	E6	0.00

(#) = Out of Range