

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082322\
 Data File : PR056094.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2022 14:45
 Operator : AJ\MA
 Sample : N4292-01
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 TR-04-082222

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/25/2022
 Supervised By :Ankita Jodhani 08/25/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 12:29:47 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:39:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.628	2.899	91163170	38438616	21.365	19.144
2) SA Decachlor...	9.519	8.111	39186854	36841123	22.039m	20.253m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082322\
Data File : PR056094.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2022 14:45
Operator : AJ\MA
Sample : N4292-01
Misc :
ALS Vial : 15 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
TR-04-082222


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/25/2022
Supervised By :Ankita Jodhani 08/25/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 12:29:47 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:39:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.628 min
Delta R.T.: -0.008 min
Response: 91163170
Conc: 21.36 ng/ml

Instrument :
ECD_R
ClientSampleId :
TR-04-082222


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/25/2022
Supervised By :Ankita Jodhani 08/25/2022




#1 Tetrachloro-m-xylene

R.T.: 2.899 min
Delta R.T.: -0.008 min
Response: 38438616
Conc: 19.14 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.519 min
Delta R.T.: -0.010 min
Response: 39186854
Conc: 22.04 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.111 min
Delta R.T.: -0.011 min
Response: 36841123
Conc: 20.25 ng/ml m