

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082421\
 Data File : PR051741.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Aug 2021 22:14
 Operator : DD\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12485245

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 04:13:47 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082421CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 25 04:10:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.626	3.805	592.9E6	690.2E6	78.866	79.139
2) SA Decachlor...	10.498	8.826	1753.4E6	1439.9E6	161.544	160.969
Target Compounds						
21) L5 AR-1248-1	5.809	4.890	276.7E6	280.7E6	1535.927	1551.296
22) L5 AR-1248-2	6.081	5.126	381.1E6	364.9E6	1516.731	1535.991
23) L5 AR-1248-3	6.288	5.168	469.9E6	391.2E6	1544.802	1539.339
24) L5 AR-1248-4	6.694	5.339	610.1E6	489.2E6	1556.000	1557.540
25) L5 AR-1248-5	6.732	5.731	595.6E6	530.9E6	1558.815	1572.439

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082421\
Data File : PR051741.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Aug 2021 22:14
Operator : DD\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :
AR12485245

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 04:13:47 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082421CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 25 04:10:51 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

