

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082720\
 Data File : PR046893.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Aug 2020 15:30
 Operator : DD\AJ
 Sample : AR1221CCC400
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 05:49:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1221PR082720CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 27 13:58:14 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.727	3.905	28556351	79946065	23.620	23.792
2)	SA Decachlor...	10.661	9.005	50850568	478.9E6	40.626	40.640
Target Compounds							
8)	L2 AR-1221-1	4.936	4.120	6622227	17397495	470.672	461.549
9)	L2 AR-1221-2	5.024	4.206	4536561	12732862	442.136	459.304
10)	L2 AR-1221-3	5.102	4.284	16494684	48588508	474.781	486.905



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082720\
Data File : PR046893.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Aug 2020 15:30
Operator : DD\AJ
Sample : AR1221CCC400
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 05:49:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1221PR082720CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 27 13:58:14 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#8 AR-1221-1


R.T.: 4.936 min
Delta R.T.: 0.000 min
Response: 6622227
Conc: 470.67 ng/ml

Instrument :
ECD_R
ClientSampleId :




#8 AR-1221-1

R.T.: 4.120 min
Delta R.T.: 0.000 min
Response: 17397495
Conc: 461.55 ng/ml




#9 AR-1221-2

R.T.: 5.024 min
Delta R.T.: 0.000 min
Response: 4536561
Conc: 442.14 ng/ml



#9 AR-1221-2


R.T.: 4.206 min
Delta R.T.: 0.000 min
Response: 12732862
Conc: 459.30 ng/ml



#10 AR-1221-3

R.T.: 5.102 min
Delta R.T.: 0.000 min
Response: 16494684
Conc: 474.78 ng/ml

Instrument :
ECD_R
ClientSampleId :



#10 AR-1221-3

R.T.: 4.284 min
Delta R.T.: 0.000 min
Response: 48588508
Conc: 486.91 ng/ml