

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082722\
 Data File : PR056204.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Aug 2022 21:20
 Operator : AJ\MA
 Sample : PB147304BL
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 PB147304BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/29/2022
 Supervised By :Ankita Jodhani 08/29/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 22:39:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:39:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.626	2.897	93024660	37050123	21.801	18.453
2) SA Decachlor...	9.514	8.106	40126165	34482631	22.568m	18.956

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR082722\
Data File : PR056204.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Aug 2022 21:20
Operator : AJ\MA
Sample : PB147304BL
Misc :
ALS Vial : 27 Sample Multiplier: 1



Instrument :
ECD_R
ClientSampleId :
PB147304BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/29/2022
Supervised By :Ankita Jodhani 08/29/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 22:39:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:39:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.626 min
Delta R.T.: -0.010 min
Response: 93024660
Conc: 21.80 ng/ml

Instrument :
ECD_R
ClientSampleId :
PB147304BL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/29/2022
Supervised By :Ankita Jodhani 08/29/2022




#1 Tetrachloro-m-xylene

R.T.: 2.897 min
Delta R.T.: -0.009 min
Response: 37050123
Conc: 18.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.514 min
Delta R.T.: -0.014 min
Response: 40126165
Conc: 22.57 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.106 min
Delta R.T.: -0.009 min
Response: 34482631
Conc: 18.96 ng/ml