

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR083123\
 Data File : PR062918.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2023 11:29
 Operator : YP\AJ
 Sample : 04175-20
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 EZYD3

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/06/2023
 Supervised By :Ankita Jodhani 09/06/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 16:00:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082823CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 29 05:38:44 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.680	2.951	32985147	50789218	14.313	16.755
2) SA Decachlor...	9.609	8.227	43912326	84159423	27.482	22.596
Target Compounds						
26) L6 AR-1254-1	5.821	4.912	4438777	7228651	49.250	38.617
27) L6 AR-1254-2	6.057	5.072	5347469	4414297	38.508	26.614 #
28) L6 AR-1254-3	6.451	5.495	6980217	11300155	49.737	41.619m
29) L6 AR-1254-4	6.760	5.742	3004917	3627174	31.639m	20.868 #
30) L6 AR-1254-5	7.217	6.186	8342481	18753259	80.421	73.431

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR083123\
Data File : PR062918.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2023 11:29
Operator : YP\AJ
Sample : 04175-20
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
EZYD3

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/06/2023
Supervised By :Ankita Jodhani 09/06/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 16:00:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR082823CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 29 05:38:44 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

