

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090120\
 Data File : PR046938.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Sep 2020 10:05
 Operator : DD\AJ
 Sample : AR1232ICC1600
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR1232501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 10:17:22 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1232PR090120CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Sep 01 09:52:30 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.727	3.917	124.1E6	406.3E6	77.906	86.416
2) SA Decachlor...	10.696	9.055	214.2E6	1984.9E6	130.061	153.746
Target Compounds						
11) L3 AR-1232-1	5.103	4.298	42907549	167.4E6	1468.124	1769.042
12) L3 AR-1232-2	5.640	5.042	22133873	190.6E6	1597.127	1764.163
13) L3 AR-1232-3	5.932	5.220	44984859	106.5E6	1560.320	1717.186
14) L3 AR-1232-4	6.096	5.305	22605645	68613615	1613.715	1699.290
15) L3 AR-1232-5	6.183	5.479	17377074	84191136	1619.821	1616.930

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090120\
Data File : PR046938.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Sep 2020 10:05
Operator : DD\AJ
Sample : AR1232ICC1600
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR1232501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 10:17:22 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1232PR090120CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Sep 01 09:52:30 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

