

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
 Data File : PR047021.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2020 19:13
 Operator : DD\AJ
 Sample : MDL-PCB-BL-S-3
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 08:01:22 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR090220CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 07:56:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.768	3.896	23853663	84471350	20.756	20.657
2) SA Decachlor...	10.730	8.995	72967738	552.6E6	43.982	44.589
Target Compounds						
26) L6 AR-1254-1	6.808	5.825	419525	7787435	6.882	29.491 #
27) L6 AR-1254-2	6.999	0.000	148302	0	1.553	N.D. #
28) L6 AR-1254-3	7.409	0.000	479338	0	4.644	N.D. #
29) L6 AR-1254-4	7.692	0.000	120639	0	1.503	N.D. #
30) L6 AR-1254-5	8.037f	0.000	554768	0	6.245	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
Data File : PR047021.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2020 19:13
Operator : DD\AJ
Sample : MDL-PCB-BL-S-3
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 08:01:22 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR090220CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 07:56:39 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 6.808 min
Delta R.T.: 0.000 min
Response: 419525
Conc: 6.88 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 5.825 min
Delta R.T.: 0.025 min
Response: 7787435
Conc: 29.49 ng/ml




#27 AR-1254-2

R.T.: 6.999 min
Delta R.T.: -0.027 min
Response: 148302
Conc: 1.55 ng/ml



#27 AR-1254-2


R.T.: 0.000 min
Exp R.T.: 5.948 min
Response: 0
Conc: N.D.



#28 AR-1254-3


R.T.: 7.409 min
Delta R.T.: 0.013 min
Response: 479338
Conc: 4.64 ng/ml

Instrument :
ECD_R
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.356 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 7.692 min
Delta R.T.: 0.011 min
Response: 120639
Conc: 1.50 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.585 min
Response: 0
Conc: N.D.



#30 AR-1254-5

R.T.: 8.037 min
Delta R.T.: -0.066 min
Response: 554768
Conc: 6.25 ng/ml

Instrument :
ECD_R
ClientSampleId :



#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 7.006 min
Response: 0
Conc: N.D.