

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
 Data File : PR047029.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2020 21:10
 Operator : DD\AJ
 Sample : MDL-PCB-BL-W-3
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 08:04:50 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR090220CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 07:56:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.771	3.896	25936546	91432398	22.569	22.359
2) SA Decachlor...	10.732	8.996	74549723	572.6E6	44.936	46.200
Target Compounds						
26) L6 AR-1254-1	6.813	0.000	200172	0	3.284	N.D. #
28) L6 AR-1254-3	7.416	0.000	217747	0	2.109	N.D. #
29) L6 AR-1254-4	7.668	0.000	20382	0	0.254	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
Data File : PR047029.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2020 21:10
Operator : DD\AJ
Sample : MDL-PCB-BL-W-3
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 08:04:50 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR090220CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 07:56:39 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#26 AR-1254-1


R.T.: 6.813 min
Delta R.T.: 0.005 min
Response: 200172
Conc: 3.28 ng/ml

Instrument :
ECD_R
ClientSampleId :




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 5.801 min
Response: 0
Conc: N.D.




#28 AR-1254-3

R.T.: 7.416 min
Delta R.T.: 0.020 min
Response: 217747
Conc: 2.11 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.356 min
Response: 0
Conc: N.D.



#29 AR-1254-4

R.T.: 7.668 min
Delta R.T.: -0.013 min
Response: 20382
Conc: 0.25 ng/ml

Instrument :
ECD_R
ClientSampleId :



#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.585 min
Response: 0
Conc: N.D.