

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
 Data File : PR047001.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2020 14:24
 Operator : DD\AJ
 Sample : MDL-AR1248-S-6
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1248-S-6

Manual Integrations
 APPROVED

Ankita
 9/3/2020 10:44:50 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 07:01:27 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1248PR090220CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Sep 02 12:28:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.756	3.895	19772344	71972903	19.607	16.757
2) SA Decachlor...	10.718	8.997	58031368	453.6E6	38.762	37.683
Target Compounds						
21) L5 AR-1248-1	5.938	4.990	858456	2544102	31.749	26.555m
22) L5 AR-1248-2	6.211	5.230	1775568	2796537	42.497	24.647 #
23) L5 AR-1248-3	6.419	5.273	1012083	2709008	22.145	20.882
24) L5 AR-1248-4	6.824	5.447	1463564	4197328	27.135	24.334
25) L5 AR-1248-5	6.864	5.842	1461650	5179690	28.632	24.876

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
Data File : PR047001.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2020 14:24
Operator : DD\AJ
Sample : MDL-AR1248-S-6
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1248-S-6

Manual Integrations
APPROVED

Ankita
9/3/2020 10:44:50 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 07:01:27 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1248PR090220CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Sep 02 12:28:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

