

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
 Data File : PR047033.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Sep 2020 22:08
 Operator : DD\AJ
 Sample : MDL-AR1254-W-4
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 MDL-AR1254-W-4

Manual Integrations
 APPROVED

Ankita
 9/4/2020 2:36:55 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 03 08:27:46 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR090220CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 03 07:56:39 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.767	3.896	16878383	59008455	14.687	14.430
2) SA Decachlor...	10.727	8.992	53325907	396.7E6	32.143	32.011
Target Compounds						
26) L6 AR-1254-1	6.807	5.800	1452873	4590304	23.833	17.383 #
27) L6 AR-1254-2	7.023	5.948	2077912	4387252	21.756	19.742
28) L6 AR-1254-3	7.393	6.355	2480045	9738404	24.026	20.146
29) L6 AR-1254-4	7.679	6.583	1836328	7654760	22.872	19.895m
30) L6 AR-1254-5	8.100	7.003	2395114	6679364	26.964	11.734 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR090220\
Data File : PR047033.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Sep 2020 22:08
Operator : DD\AJ
Sample : MDL-AR1254-W-4
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
MDL-AR1254-W-4

Manual Integrations
APPROVED

Ankita
9/4/2020 2:36:55 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 03 08:27:46 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\AR1254PR090220CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 03 07:56:39 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

